Validation of Fluorescent SSCP Analysis for Sensitive Detection of p53 Mutations

BioTechniques 28:986-992 (May 2000)

**L. Moore, T. Godfrey, C. Eng,
A. Smith, R. Ho and
F.M. Waldman**

University of California Berkeley,
School of Public Health,
Berkeley, CA, USA

INTRODUCTION

The p53 tumor suppressor protein is involved in many important cellular processes that include apoptosis, gene transcription, DNA repair, genomic stability and cell cycle control (2). Therefore, it is not surprising that mutation of the p53 gene is the most common genetic alteration in human cancers. Specific patterns of p53 mutations are thought to result from different mechanisms, and mutations in different protein domains are associated with varying functional alterations (5). By characterizing the frequency, type and position of mutations in the p53 gene, inferences may be made regarding potential sources of mutation in the environment.

The p53 gene contains 11 exons and 393 amino acids (7). Of these, exons 5–8 contain “hot spots” for mutation. Such preferred sites reside in evolutionarily conserved regions of the gene, depending on their functional importance (3). DNA sequencing is required for precise identification of mutations, but this can be labor intensive and expensive. To overcome these problems, screening tests are frequently used to preselect samples with possible mutations for subsequent sequencing.

While new methods for mutation detection are constantly being developed, single strand conformation polymorphism (SSCP) (9) is still the most commonly cited technique. The visualization of DNA strands is usually carried out by isotopic labeling of PCR primers, followed by autoradiography or staining the gel with agents such as ethidium bromide or silver. None of these visualization methods is capable of distinguishing between the two DNA strands, and the resulting data are often hard to interpret, especially when complicated by more than one stable conformation of one or both strands.

Recently, several authors have used

fluorescent labeling to uniquely identify sense and antisense DNA strands (1,4,8,10). Combined with automated DNA sequencers for accurate temperature control and multicolor fluorescence detection, these features have the potential to significantly improve the 70%–90% mutation detection rates previously reported (8,9). The system also includes internal size standards in each lane and thus allows small mobility shifts to be reliably detected.

Given the importance of p53 mutation detection, any improvements of current methods will be useful, especially in screening multiple tumor samples for clinical and molecular epidemiological studies. Here, we report the optimization of a fluorescence-based SSCP-PCR method that overcomes most of the difficulties associated with conventional SSCP procedures and allows rapid and accurate mutation screening of p53 exons 5–8. Moreover, sensitivity and specificity (the percentage of true mutants/all mutant samples and the percentage of true normals/all normal samples, respectively) of SSCP screening—the critical factors that strongly influence interpretation of p53 mutation studies—have been measured using previously characterized (6) mutant cell lines and normal tissues. Mixing experiments using heterozygous mutant and normal DNAs were also conducted to determine the limits of mutation detection in heterogeneous samples.

MATERIALS AND METHODS

Cell Lines

Representing unique mutations and amino acid substitutions throughout exons 5–8, 21 cell lines were identified in the International Agency for Cancer Research’s p53 database (reference 6, which is also available electronically at

ABSTRACT

We have developed a fluorescence-based single strand conformation polymorphism (SSCP) method that offers fast and sensitive screening for mutations in exons 5–8 of the human p53 gene. The method uses an ABI 377 DNA sequencer for unique color detection of each strand, plus accurate alignment of lanes for better detection of mobility shifts. To validate the method, 21 cell lines with reported mutations in p53 exons 5–8 were analyzed by SSCP using various gel conditions. The sensitivity for mutation detection was 95% for all cell lines studied, and no false positives were seen in 10 normal DNA samples for all four exons. Experiments mixing known amounts of tumor and normal DNA showed that mutations were detected even when tumor DNA was mixed with 80% normal DNA. Fluorescent SSCP analysis using the ABI sequencer is a useful tool in cancer research, where screening large numbers of samples for p53 mutations is desired.

Table 1. SSCP for Mutant Cell Lines using Optimal Conditions

Exon	Cell Line	Codon	Mutation	Sense Strand (mobility shifts)	Antisense Strand (mobility shifts)
5 (276 bp ^a)	BT20	132	AAG-CAG	- ^a (0) ^b	+ ^c (-486)
5	ST486	158	CGC-CAC	+ (+74)	+ (157)
5	CEM	175	CGC-CAC (Heterozygote)	+ (-122)	+ (-905)
5	LAN1	182	TGC-TGA	+ (-187)	+ (-52)
5	SW1116	159	GCC-GAC	+ (+449)	+ (-397)
5	SKBR3	175	CGC-CAC	+ (-112)	+ (-927)
Sensitivity				6/6 = 100%	
6 (165 bp)	Raji	213	CGA-CAA (Heterozygote)	+ (-65)	- (0)
6	Calu6	196	CGA-TGA	+ (-258)	+ (-70)
6	DU145	223	CCT-CTT	+ (+132)	- (0)
6	T47D	194	CTT-TTT	+ (-63)	+ (+47)
6	SKLU1	193	CAT-CGT	- (0)	- (0)
Sensitivity				4/5 = 80%	
7 (182 bp)	Raji	234	TAC-CAC	- (0)	+ (-74)
7	RD	248	CGG-TGG	+ (-182)	+ (+154)
7	CEM	248	CGG-CAG	+ (+217)	+ (-102)
7	BT549	249	AGG-AGC	+ (-223)	+ (-284)
7	Ramos	254	ATC-GAC	+ (+891)	+ (+753)
7	DAOY	242	TGC-TTC	+ (+218)	+ (+167)
Sensitivity				6/6 = 100%	
8 (237 bp)	MDA231	280	AGA-AAA	+ (+77)	- (0)
8	BT474	285	GAG-AAG	+ (+103)	+ (-47)
8	DU145	273	CGT-CAT (Heterozygote)	+ (+73)	+ (+68)
8	PANC1	274	GTT-TTT ^d	- (0)	- (0)
Sensitivity				3/3 = 100%	
Sensitivity for all Exons				19/20 = 95%	
^a - indicates no gel shift seen					
^b indicates the number of mobility shifts (in scans) and direction with respect to wild-type					
^c + indicates a gel shift was present,					
^d this mutation was not confirmed by sequencing					

www.ftp.embl-heidelberg.de/pub/data-bases/p53) (Table 1). Cell lines, readily available from the cell culture facility at the University of California at San Francisco (UCSF) and containing different types of mutations in exons 5–8, were used. Two normal DNA samples were obtained from healthy donors' blood, and eight samples were from frozen normal lung tissue (supplied by Dr. David Jablons at UCSF). DNA was

extracted using the Wizard® genomic DNA purification kit according to the manufacturer's instructions (Promega, Madison, WI, USA).

PCR

PCR was carried out in a GeneAmp® PCR System 9700 thermal cycler (PE Biosystems, Foster City, CA, USA) in 25-μL reaction volumes. Each

PCR contained 2.5 mM MgCl₂, 0.2 mM each dNTP, 0.625 U *Taq* Gold® polymerase, 1× PCR Buffer II (both from PE Biosystems), 0.4 μM of sense and antisense fluorescently labeled primers (Table 2) and 20–40 ng genomic DNA. PCR cycling conditions were as follows: 95°C for 12 min followed by 32 cycles at 95°C for 15 s, 58°C for 15 s, 72°C for 30 s and a final extension at 72°C for 10 min. Table 2 shows the

Research Report

Table 2. Primer Sequences Used for SSCP and Sequencing

Exon	SSCP PCR Primers	Sequencing PCR Primers ^a	Sequencing Primers
E5 forward	FAM-CCA GTT GCT TTA TCT GTT CA	TGT CGT CTC TCC AGC CCC AG	TGT CGT CTC TCC AGC CCC AG
E5 reverse	TET- TGT CGT CTC TCC AGC CCC AG	CCA GTT GCT TTA TCT GTT CA	CCA GTT GCT TTA TCT GTT CA
E6 forward	FAM- GCT GCT CAG ATA GCG ATG GTG	M13(f)-CTC TGA TTC CTC ACT GAT TG	TGT AAA ACG ACG GCC AGT
E6 reverse	TET-GCC ACT GAC AAC CAC CCT TAA C	M13(r)- CAG AGA CCC CAG TTG CAAAAC	CAG GAA ACA GCT ATG ACC
E7 forward	FAM-CCA AGG CGC ACT GGC CTC	M13(f)-CCA AGG CGC ACT GGC CTC	TGT AAA ACG ACG GCC AGT
E7 reverse	TET-AGT GTG CAG GGT GGC AAG	M13(r)-AGT GTG CAG GGT GGC AAG	CAG GAA ACA GCT ATG ACC
E8 forward	FAM- CCT TAC TGC CTC TTG CTT C	M13(f)-CCT TAC TGC CTC TTG CTT C	TGT AAA ACG ACG GCC AGT
E8 reverse	TET-TGA ATC TGA GGC ATA ACT GC	M13(r)-TGA ATC TGA GGC ATA ACTGC	CAG GAA ACA GCT ATG ACC

^aM13(f)-forward, TGTAACGACGGCCAGT; M13(r)-reverse, CAGGAAACAGCTATGACC

PCR primers and fluorescent labels used for each exon. All primers were purchased from Life Technologies (Rockville, MD, USA).

SSCP

A 5.5-μL loading mixture was prepared with 1 μL PCR product, 0.5 μL of GeneScan® 500 size standards (PE Biosystems), 0.5 μL loading dye (PE Biosystems), 0.5 μL 100 mM NaOH and 3 μL deionized formamide. The loading mixture was denatured at 95°C for 3 min and then placed immediately on ice. Then, 0.7–1 μL of this sample mixture were loaded into each well of the SSCP gel. Multiple normals were electrophoresed on each gel (every 5–7 lanes) to aid data analysis in case of uneven gel performance. In addition, only the center lanes were used (30 of 36 or 40 of a 48-lane comb) to avoid uneven temperature control in the outer lanes. To determine positivity, all positive shifts reported in this paper were > 2(± SD) from the mean of the wild-type DNA samples electrophoresed on the same gel.

SSCP gels were cast using 36-cm plates (PE Biosystems) and 36- or 48-

well shark’s-tooth combs. All gels consisted of 0.5× mutation detection enhancement (MDE) gel (BMA, Rockland, ME, USA) with 2.5%, 5% or 10% glycerol (Sigma, St. Louis, MO, USA) and 1× TBE (Life Technologies). The gels were polymerized with 300 μL 10% ammonium persulphate and 37.5 μL TEMED per 60 mL gel solution. The gels were then mounted on the pre-cooled ABI 377 sequencer, and no pre-run was used before loading. Electrophoresis was performed at 15°C, 20°C and 25°C. Gel cooling was provided by a VWR 1190A refrigerated water bath (VWR Scientific, West Chester, PA, USA) connected to the inlet/outlet tubes of the ABI 377 sequencer. The 2400C chiller module was used in the ABI collection software to override the instrument’s pump and temperature control systems that were provided by the water bath instead. Gel temperature was monitored in the software status window. All gels were run at 60 W constant power with nonlimiting voltage (4000 V) and current (60 mA).

Data Analysis

SSCP results were analyzed using

GeneScan software with analysis procedures and settings as described by the manufacturer’s ABI SSCP user guide. The GeneScan software allows test DNA data to be overlaid with data from a normal sample on the same gel to make visualization of mobility shifts easier. The sense and antisense strands were scored separately for shifts.

Sequencing

Fluorescent DNA sequencing was carried out using BigDye™ dye terminator sequencing chemistry (PE Biosystems). Primers used are shown in Table 2. Before sequencing, PCR products were purified using the PCR product presequencing kit (Amersham Pharmacia Biotech, Piscataway, NJ, USA) to remove excess primers and nucleotides. All exons were sequenced in both forward and reverse directions to confirm the presence or absence of mutations.

RESULTS

SSCP was performed on 21 cell lines with previously reported mutations (6). Table 1 shows SSCP results for the op-

timum conditions (0.5× MDE, 2.5% glycerol, 25°C). Both the sense (FAM) and antisense (TET) strands were shifted more frequently using these conditions. Overall, the sensitivity of SSCP using these conditions was 95%, and there were no false positives detected in the normal DNA tested. Examples of SSCP results for each exon, a comparison of mutant cell line to normal PCR product, are presented in Figure 1. When the same samples were run at 20°C with 2.5% glycerol, we detected 19/21 or 90.5% of mutations and were unable to detect two mutations in exon

6, codon 223 of cell line DU145 and codon 213 of cell line Raji.

When the glycerol content was increased to 5% at 20°C, 14/16 or 87.5% of mutations were detected, and we were unable to detect the mutations in exon 5, codon 158 of cell line ST486 and exon 8 and codon 273 of cell line DU145. When the temperature was increased to 25°C, 14/15 or 93% of mutations were detected. Using this condition, it was impossible to detect the exon 5 codon 182 mutation described in cell line LAN1. The worst results were obtained when 10% glycerol was used. At 25°C,

Figure 1. Fluorescent SSCP detection in cell lines. In each panel, each cell line DNA was amplified with exon-specific primers, each labeled with either FAM (sense) or TET (antisense). A normal profile was overlaid with the test profile and aligned using the internal size standards (small red peaks). In each of the panels, the normal-sense strand is labeled red-orange, and the normal-antisense strand is labeled brown. The cell line-sense strand is labeled blue, and the cell line-antisense strand is labeled green. (A) Heterozygous mutation in CEM cell line exon 5. Sense (blue) and antisense (green) strands show both shifted and unshifted peaks. (B) Homozygous mutation in T47D cell line exon 6. Shifts can be seen in both the sense (blue) and antisense (green) strands. (C) Heterozygous mutation in RAJI cell line exon 7. A shift can be detected in the antisense (green) but not the sense (blue) strand. (D) Homozygous mutation in BT474 cell line exon 8. Multiple peaks are seen for both sense and antisense strands as a result of multiple stable conformations. All peaks (sense and antisense) are shifted relative to normal control.

Research Report

14/16 or 75% of mutations were detected. Exon 6 failed in all cases with 5% or 10% glycerol. Therefore, the results are given for only 15 or 16 cell lines.

A mutation in cell line PANC1 was not detected with any of the SSCP gel conditions. DNA sequencing showed no evidence of the mutation that was reported in the database. Sequencing of all remaining cell lines confirmed the previously reported mutations.

Mixing experiments were used to define the sensitivity of SSCP analysis in a pre-PCR mixture of mutated formalin-fixed tumor and normal DNA (mixed before PCR amplification). An abnormally migrating peak was clearly seen, even in the presence of 80% and possibly even as low as 90% normal DNA (arrows, Figure 2, panels D and E).

DISCUSSION

There are numerous advantages for using multicolor fluorescent SSCP analysis compared to using radioactive labeling methods. One is that isotopes are not used. Second, all analysis is done using a computer imaging system, eliminating the need for post-electrophoresis gel handling and achieving better resolution of mobility shifts to align data from lane to lane. Third, the use of dual colors to label the individual strands allows overlapping bands to be discriminated and also allows easier interpretation of band shifts associated with specific mutations.

Recent advances in sequencing technology bring into question the general utility of SSCP analysis. However, we have found that preliminary screening for mutations with SSCP is practical and efficient, especially when conducting large studies. SSCP screening is faster and less expensive than sequencing, especially if the percentage of mutant samples is low. Furthermore, SSCP can determine which exon is mutated, and only this exon needs to be sequenced for confirmation, thus reducing the amount of sequencing required. If the sequencing reaction has low background and so is easy to interpret, then SSCP followed by sequencing in only one direction can be used. However, if the sequencing reaction contains a high background or if the sample is heterogeneous, it is opti-

mal to use SSCP followed by bidirectional sequencing. Alternatively, SSCP is a sensitive method to confirm sequencing abnormalities. Similarly, SSCP could be used to confirm the presence or absence of heterozygotes initially identified by sequencing but obscured by a high background. Background in sequencing frequently results from a poor DNA template, a common occurrence when using DNA from formalin-fixed paraffin sections.

A previous study by O'Connell et al. (8) tested the MDE gel but found that

optimal results were obtained with a 12% (w/v) acrylamide 50:1 plus 10% (v/v) glycerol gel run at 20°C. They noted small effects for the Daudi, exon 5 codon 153 and RD-ES, exon 8 codon 273 mutations, but large mobility shifts for other mutations in exons 7 and 8, some of which were nearly too large to detect after 14 h. With the addition of 15% urea to the MDE gel, a reduction was seen in the unusually large mobility shifts and samples not previously detectable in the absence of urea. In our study, we found that the MDE gel ran

Figure 2. Mixing experiment to determine the sensitivity of mutation detection in the presence of normal DNA. (A) 100% tumor (T) and 100% normal (N) profiles overlaid. (B) 40% tumor and 60% normal. (C) 30% tumor and 70% normal. (D) 20% tumor and 80% normal (E) 10% tumor and 90% normal. (F) 100% normal. Arrows indicate the tumor peaks visible even in the presence of 80% normal DNA.

Research Report

more consistently, offered greater resolution than acrylamide and was therefore more likely to detect mobility shifts. When glycerol levels were 5% or greater, we found that both consistency and resolution decreased. The worst results were obtained when 10% glycerol was used.

O'Connell et al. (8) also found considerable variability in the detection sensitivity of the seven p53 mutations analyzed using different gel conditions. In our study, 21 rather than seven cell lines, with mutations throughout exon 5–8 and 10 normal controls, were analyzed. The sensitivity of mutation detection was 95%, which is higher than the 70%–90% mutation detection rates previously reported by some groups (2,9), but lower than 100% as reported by O'Connell et al., possibly because of the large number of cell lines tested (8).

Furthermore, it was possible to detect mutant tumor DNA in the presence of 80% normal DNA. This level of sensitivity in mutation detection is important since many tumors are

heterozygous or heterogeneous for p53 mutations, and since tumor tissue is frequently contaminated with normal cells. In such samples, the presence of a mutation would be difficult to identify by sequencing alone (especially in the presence of any sort of background), but the identification of a mutation using both fluorescent SSCP and sequencing is substantially improved. However, it should be noted that fluorescent SSCP does not allow isolation or enrichment of possible mutants for sequence analysis, which is possible with conventional SSCP.

For the cell lines analyzed here, all positive shifts reported were greater than 2 SD from the mean of the wild-type samples run on the same gel. However, when screening unknown samples for mutations, we preferred to use the SD of the normal samples only as a guide, with the final decision being made subjectively by aligning the test samples in only the closest lanes. This is because normal peaks often show a nonrandom and gradual shift from one side of the gel to the other, probably caused by slight variations in temperature or gel thickness. By aligning each test case with the two closest normal samples, it was possible to control for such variations across the gel. In our studies, because all suspected mutants are eventually sequenced, this methodology does not increase the number of false positives but slightly increases the amount of required sequencing.

In conclusion, we have further validated a new fluorescence-based SSCP method that allows fast and sensitive screening for mutations in exons 5–8 of the human p53 gene. The method uses an ABI 377 sequencer for unique color labeling of each strand and accurate alignment of lanes for better detection of mobility shifts. Because this method has greater sensitivity and specificity than other methods previously described, it is a useful tool in areas of cancer research where screening large numbers of samples for p53 mutations is desired.

ACKNOWLEDGMENTS

Support for this work was provided by grant nos. P30-ES01896 and P42-ES04705, both from the National Insti-

tute of Environmental Health Science, and grants from the UCSF Cancer Center. L.M. is the recipient of American Cancer Society fellowship no. PF4440.

REFERENCES

1. **Boutin, P., E.H. Hani, F. Vasseur, C. Roch, B. Bailleul, J. Hager and P. Frognel.** 1997. Automated fluorescent-based screening for mutation by SSCP: use of universal M13 dye primers for labeling and detection. *BioTechniques* 23:358-362.
2. **Greenblatt, M.S., W.P. Bennett, M. Hollstein and C.C. Harris.** 1994. Mutations in the p53 tumor suppressor gene: clues to cancer etiology and molecular pathogenesis. *Cancer Res.* 54:4855-4878.
3. **Harris, C.C.** Structure and function of the p53 tumor suppressor gene: clues for rational cancer therapeutic strategies. 1996. *J. Natl. Cancer Inst.* 88:1442-1455.
4. **Hayashi, K.** 1999. Recent enhancements in SSCP. *Genetic Anal.* 14:193-196.
5. **Hollstein, M., D. Sidransky, B. Vogelstein and C.C. Harris.** 1991. p53 in human cancers. *Science* 253:49-53.
6. **Hollstein, M., K. Rice, M.S. Greenblatt, T. Soussi, R. Fuchs, T. Sorlie, E. Hovig, B. Smith-Sorensen et al.** 1994. Database of p53 gene somatic mutations in human tumors and cell lines. *Nucleic Acids Res.* 22:3551-3555.
7. **Lamb, P. and L. Crawford.** 1986. Characterization of the human p53 gene. *Mol. Cell. Biol.* 6:1379-1385.
8. **O'Connell, C.D., J. Tian, A. Juhasz, H.M. Wenz and D.H. Atha.** 1998. Development of standard reference materials for diagnosis of p53 mutations: analysis by slab gel single strand conformation polymorphism. *Electrophoresis* 19:164-171.
9. **Orita, M., H. Iwahana, H. Kanazawa, K. Hayashi and T. Sekiya.** 1989. Detection of polymorphisms of human DNA by gel electrophoresis as single-strand conformation polymorphisms. *Proc. Natl. Acad. Sci. USA* 86:2766-2770.
10. **Wenz, H.M., S. Ramachandra, C.D. O'Connell and D.H. Atha.** 1998. Identification of known p53 point mutations by capillary electrophoresis using unique mobility profiles in a blinded study. *Mutat. Res.* 382:121-132.

Received 16 July 1999; accepted 7 December 1999.

Address Correspondence to:

Dr. Lee Moore
University of California Berkeley
School of Public Health
140 Warren Hall
Berkeley, CA 94720-7360 USA
Internet: lmoore@cc.ucsf.edu