Benchmarks

products, respectively. The fluorescence from tubes containing specific products also could be visualized by eye using simple UV transillumination. As expected, the PCR negative control (no template DNA) did not show any significant change in fluorescence, even after 40 cycles of amplification.

The protocol outlined in this work takes advantage of the fact that the synthesis cost of single labeling (\$180 US for 25 nmol probe set in this test) is lower than that for dual-labeled probes (\$250–\$400 US for 10–20 nmol dual-labeled probes). This cost difference could benefit researchers engaged in assay development on a limited budget. For example, more asymmetric probe sets could be evaluated on a dollar-per-dollar basis, and a single quencher probe could be used with several alternative reporter probes in assay optimization. Further improvements and applications of this design, specifically for routine quality control and sanitary monitoring (in endpoint detection format) in food or other related industries, are underway in our laboratory.

REFERENCES

1. Bassler, H.A., S.J.A. Flood, K.J. Livak, J. Marmaro, R. Knorr and C.A. Batt. 1995. Use of fluorogenic probe in a PCR-based assay for the detection of *Listeria monocytogenes*. *Appl. Environ. Microbiol.* 61:3724–3728.
2. Gibson, N.J., C.R. Newton and S. Little. 1997. A colorimetric assay for phosphate to measure amplicon accumulation in polymerase chain reaction. *Anal. Biochem.* 254:18–22.
3. Higuchi, R., D. Dollinger, P.S. Walsh and R. Griffith. 1992. Simultaneously amplification and detection of specific DNA sequences. *BioTechnology* 10:413–417.
4. Holland, P.M., R.D. Abramson, R. Watson and D.H. Gelfand. 1991. Detection of specific polymerase chain reaction product by utilization the 5'→3' exonuclease activity of *Thermus aquaticus* DNA polymerase. *Proc. Natl. Acad. Sci. USA* 88:7276–7280.
5. Lee, L.G., C.R. Connell and W. Bloch. 1993. Allelic discrimination by nick-translation PCR with fluorogenic probes. *Nucleic Acids Res.* 21:3761–3766.
6. Livak, K.J., S.J.A. Flood, J. Marmaro, W. Giusti and K. Deetz. 1995. Oligonucleotides with fluorescent dyes at opposite ends provide a quenched probe system useful for detecting PCR product and nucleic acid hybridization. *PCR Methods Appl.* 4:357–362.
7. Morrison, L.E., T.C. Halder and L.M. Stols. 1989. Solution-phase detection of polynucleotides using interacting fluorescent labels and competitive hybridization. *Anal. Biochem.* 183:231–244.
8. Nazarenko, I.A., S.K. Bhatnagar and R.J. Hohman. 1997. A closed tube format for amplification and detection of DNA based on energy transfer. *Nucleic Acids Res.* 25:2516–2521.
9. Tyagi, S and F.R. Kramer. 1996 Molecular beacons: probes that fluoresce upon hybridization. *Nat. Biotechnol.* 14:303–308.
10. Wang, C.N.J., K.Y. Wu and H.-T. Wang. 1995. Quantitative PCR using the AmpliSensor assay, p. 193–202. In C.W. Diffenbach and G.S. Dveksler (Eds.), *PCR Primer: A Laboratory Manual*. CSH Laboratory Press. Cold Spring Harbor, NY.
11. Whitcombe, D., J. Theaker, S.P. Guy, T. Brown and S. Little. 1999. Detection of PCR products using self-probing amplicons and fluorescence. *Nat. Biotechnol.* 17:804–807.
12. Wittwer, C.T., M.G. Herrmann, A.A. Moss and R.P. Rasmussen. 1997. Continuous fluorescence monitoring of rapid cycle DNA amplification. *BioTechniques* 22:130–138.

This manuscript is no. FSR-00-2 in the journal series of the Department of Food Science, North Carolina State University, Raleigh, NC, USA. This study was funded by Dairy Management, Inc., through the Southeast Dairy Foods Research Center. Address correspondence to Dr. Lee-Ann Jaykus, Department of Food Science, North Carolina State University, Raleigh, NC 27695-7624, USA. e-mail: leean_jaykus@ncsu.edu

Received 25 January 2000; accepted 14 June 2000.

Kai Koo and Lee-Ann Jaykus
*North Carolina State University
Raleigh, NC, USA*

Method for Avoiding PCR-Inhibiting Contaminants when Eluting DNA from Polyacrylamide Gels

BioTechniques 29:694–696 (October 2000)

Different methods have been reported for the elution of DNA from poly-

acrylamide gels for PCR amplification. Most of these describe protocols in which the gel piece is soaked in water at different temperatures and for different incubation times followed by amplification of the eluate (3,6). Alternatively, PCR is performed after purification of the eluate by precipitation of DNA or filtration to remove contaminants (5). Also, electroelution of DNA from the gel has been described (4). We have examined the conditions for DNA elution from polyacrylamide gel pieces and arrived at a rapid method that consistently gives high yields of products after re-amplification by PCR. As an example of the procedure, elution and re-amplification of HLA heteroduplexes resolved by non-denaturing PAGE is described.

HLA-A heteroduplexes were resolved by 5% PAGE in 1× TBE as previously described (1). When comparing DNA eluted from gels into distilled water at ambient temperature overnight (3) and at 95°C for 1 h (6), we found that eluates obtained at elevated temperatures generally gave lower DNA yields after re-amplification than did DNA eluates obtained at ambient temperature. However, DNA eluates obtained at elevated temperatures amplified well after purification (data not shown) (5). The addition of blank (polyacrylamide gel slices without DNA) eluates obtained at 95°C to PCR of HLA-A exon 2 led to reduced product yields (data not shown). We suspected that PCR inhibitors might co-elute from the non-denaturing acrylamide gel with the DNA at high temperature. Since overnight incubation at ambient temperature is time consuming (3), we investigated the minimal time required for DNA elution from non-denaturing polyacrylamide gels at ambient temperature.

A gel slice (about 40 mg) containing resolved HLA-A PCR products excised from 5% polyacrylamide gels was soaked in 100 µL distilled water for 15 min at ambient temperature (22°C–25°C) in a microcentrifuge tube. The supernatant was removed, 100 µL distilled water was added and the tube was vigorously shaken for 5 s and spun for 1 min at 5000× g. Aliquots of the supernatant (5 µL) were removed at 0-, 15-, 30-, 45-, 60- and 120-min incubations at ambient temperature. After removal of the last aliquot, the gel piece

Benchmarks

was further incubated at 37°C for 15 min, spun as described above and a 5- μ L aliquot was removed. The eluates were then used as templates for PCR as described in the legend to Figure 1.

Increasing inhibition of PCR re-amplification with increasing incubation time and elevated temperature was observed (Figure 1). The best amplification yields were obtained from the supernatant and 0-min eluates, while the lowest signal was seen in the 135-min, 37°C eluate. Also, a similar gel piece processed as described above and incubated at 37°C for 15 min yielded no PCR re-

Figure 1. Inhibition of PCR re-amplification of DNA eluted from non-denaturing polyacrylamide gels at different time points (0–120 min). The reaction was comprised of 5 pmol each HLA-AF/HLA-AR primers, 1 $\times$ ammonium buffer, 0.2 mM dNTPs, 1 mM MgCl₂, 0.5 U *Taq* DNA polymerase (Biolone, London, UK) and 2.5 μ L of the respective eluates serving as templates in a total volume of 25 μ L (1,2). W = wash eluate, 135 37° = eluate from gel piece incubated at 37°C for 15 min after 2 h incubation. DNA size marker is *Eco*RI-*Bst*II digest of λ DNA.

Figure 2. Acrylamide inhibition of PCR. The HLA-A exon II was amplified from genomic DNA in the presence of acrylamide at final concentrations of 1.0%, 0.8%, 0.6%, 0.4%, 0.2%, 0.16%, 0.05%, 0.03% and 0.00%, respectively, using genomic DNA as the template. PCR conditions and temperature cycling were similar to that described in the Figure 1 legend except for the substitution of HLA-exon II reverse primer (HLA-AR) for HLA-A forward primer (HLA-AF). DNA size marker is a 1-kb Plus DNA Ladder (Life Technologies, Rockville, MD, USA).

amplification products (data not shown). These observations indicate that a PCR inhibitor might have been released from the polyacrylamide gel pieces along with eluted DNA fragments. The amount of the inhibitor released from the gel pieces appears to be dependent on incubation time and temperature.

In the absence of known PCR inhibitors in our native polyacrylamide gels, we suspected that non-polymerized acrylamide monomers might contribute to the observed inhibition because 10% of the acrylamide may remain non-polymerized after persulfate/TEMED induction of polymerization (7). Various concentrations of acrylamide were added to reactions aimed at amplifying HLA-A exon II from 10 ng human genomic DNA (Figure 2). The result shows that acrylamide at concentrations greater than 0.05% inhibited the amplification of the gene segment under the specified conditions. Provided that all the non-polymerized acrylamide molecules in a 40- μ L 5% polyacrylamide gel piece is eluted in 100 μ L water and that the eluate is diluted 1:10 in the PCR mixture, the concentration of eluted acrylamide in the PCR mixture will be approximately 0.02%. Thus, the observed inhibition of PCR by eluates from polyacrylamide gels is probably not due to eluted acrylamide monomers alone.

Our data indicate that DNA suitable for PCR amplification may be eluted from acrylamide gels rapidly and easily. The gel piece containing the DNA of interest is soaked in 100 μ L distilled water, mixed briefly and the DNA-containing supernatant removed after centrifugation at 5000 $\times$ g for 1 min. Centrifugation at 13 000 $\times$ g did not change the re-amplification yields (data not shown). We have used 0.5–5 μ L eluate as template in the PCR scheme described in the legend to Figure 1. The method is sensitive because as little as 100 pg PCR products, resolved by 4%–10% non-denaturing PAGE, have been successfully eluted and re-amplified to levels detectable by electrophoresis on a 1% agarose gel (data not shown).

In conclusion, we have shown that DNA, in sufficient amounts for PCR re-amplification, may be eluted from polyacrylamide gel slabs in less than 2 min. We have demonstrated that PCR in-

hibitors co-elute with DNA from native polyacrylamide gels in a time- and temperature-dependent fashion. Our results also demonstrate that acrylamide is a potent inhibitor of PCR, but non-polymerized acrylamide is probably not the only PCR inhibitor in the gel eluates.

REFERENCES

1. Arguello, R., A.L. Pay, A. McDermott, J. Ross, P. Dunn, H. Avakian, A.-M. Little, J. Goldman and J.A. Madrigal. 1997. Complementary strand analysis: a new approach for allelic separation in complex polyallelic genetic systems. *Nucleic Acids Res.* 25:2236–2238.
2. Cereb, N., P. Maybe, S. Lee, Y. Kong and S. Yang. 1995. Locus-specific amplification of HLA class I genes from genomic DNA: locus specific sequences in the first and third introns of HLA-A, -B and -C alleles. *Tissue Antigens* 45:1–11.
3. Eberle, M., L.A. Knapp, K.K. Iwanaga, M.J. Domanico, K. Aiyer and D.I. Watkins. 1997. HLA-B typing by allele separation followed by direct sequencing. *Tissue Antigens* 49:365–375.
4. Karupiah, N. and P.B. Kaufman. 1992. Rapid and inexpensive micro-elution of nucleic acid and protein from agarose and polyacrylamide gels. *BioTechniques* 13:366.
5. Konecny, P. and M.G. Redinbaugh. 1997. Amplification of differentially displayed PCR products isolated from untreated denaturing polyacrylamide gels. *BioTechniques* 22:240–244.
6. Raju, R., B.L. Hoppe, D. Navneetham and B.M. Conti-Fine. 1995. Rapid method for the elution and analysis of PCR products separated on high resolution acrylamide gel. *BioTechniques* 18:33–35.
7. Reghetti, P.G. and C. Gelfi. 1997. Electrophoresis gel media: the state of the art. *J. Chromatogr.* 669:63–75.

The authors thank Prof. Frode Vartdal of our institute for critically reading the manuscript. Financial support was obtained from Medinnova SF, Oslo, Norway. Address correspondence to Dr. Anne Spurkland, Institute of Immunology, The National Hospital, 0027 Oslo, Norway. e-mail: anne.spurkland@embnet.uio.no

Received 22 December 1999; accepted 27 June 2000.

Godfrey Essien Etokebe and Anne Spurkland
The National Hospital
Oslo, Norway