Benchmarks

culated. During analysis, a cut-off value is set up for the spot size at a minimum of six pixels to eliminate any small nonspecific background spots, and this value is kept constant.

Once the system is set up, a macro is written with the optimal settings. This makes the system more efficient and reproducible. It takes around 20 s to count each well of the microplate, and up to 400 spots/well can be counted quite accurately compared to the exhausting exercise of counting as many spots manually under a microscope. Occasional clusters of cells can be easily and accurately counted by using the "split clustered objects" function of the Image-Pro software. The whole system can be set up at less than one-tenth the cost of the commercially available digital quantification systems. The results are more consistent than manual counts in which the counts can vary between two investigators, especially when there are too many spots/well. It is difficult to manually count and also to differentiate between specific and nonspecific spots if there are more than 200 spots/well, and this can affect the consistency. As reported earlier (8), we also observed that, although the results of manual reading correlate with those counted by the computer, there is a greater standard deviation in manual counts and that the counts are always lower, especially in wells having more than 100 spots. For example, the mean $\pm$ standard deviation of four counts of a particular well was 200 ± 1.6 by the lens-assisted computer-based counting as compared to 188 ± 23 spots counted manually under a microscope by four individuals. Other image analysis software available in the laboratory can also be used. The National Institutes of Health (NIH) Image Version 1.55 shareware (<http://rsb.info.nih.gov/niimage>) for ELISPOT quantification can further cut down costs. This software has been used earlier by Cui and Chang (1); however, their overall set up is very complicated: the image is captured, optimized by the computer, saved in a picture file format, and then processed after improving the resolution by the help of a computer. A modified version of this setup was later reported (4) with a fully automated system that, besides being more expensive, also requires the

removal of the nitrocellulose membrane from the plates. Recently, the AlphaImager imaging system has been introduced and is commercially available (8). However, our system based on the same principle can be set up in-house at a fraction of the cost and is equally efficient. Overall, the advantages of our system over those previously published (1,4) are: (i) the direct attachment of a microscope lens to the CCD camera eliminates the requirement for a microscope; (ii) the complicated and time-consuming steps of saving each image in a picture file format followed by image resolution by the computer are also eliminated because the microscope lens magnifies and helps in transporting a highly resolved image of the spots on to the computer screen, which is instantly counted; and (iii) the automatic quantification of these spots by the Image-Pro Plus, or any other available standard imaging software, ensures reproducible and efficient quantification and can also, if required, supply information on other parameters, for example, the area of each spot (Figure 1).

REFERENCES

1. Cui, Y. and L.J. Chang. 1997. Computer-assisted, quantitative cytokine enzyme-linked immunospot analysis of human immune effector cell function. *BioTechniques* 22:1146-1149.
2. Czerkinsky, C., G. Andersson, H.P. Ekre, L.A. Nilsson, L. Klareskog, and O. Ouchterlony. 1988. Reverse ELISPOT assay for clonal analysis of cytokine production. I. Enumeration of gamma-interferon-secreting cells. *J. Immunol. Methods* 110:29-36.
3. Hanke, T., T.J. Blanchard, J. Schneider, G.S. Ogg, R. Tan, M. Becker, S.C. Gilbert, A.V. Hill, G.L. Smith, and A. McMichael. 1998. Immunogenicities of intravenous and intramuscular administrations of modified vaccinia virus Ankara-based multi-CTL epitope vaccine for human immunodeficiency virus type 1 in mice. *J. Gen. Virol.* 79:83-90.
4. Herr, W., B. Linn, N. Leister, E. Wandel, K.H. Meyer zum Buschenfelde, and T. Wolfel. 1997. The use of computer-assisted video image analysis for the quantification of CD8+ T lymphocytes producing tumor necrosis factor alpha spots in response to peptide antigens. *J. Immunol. Methods* 203:141-152.
5. McCutcheon, M., N. Wehner, A. Wensky, M. Kushner, S. Doan, L. Hsiao, P. Calabresi, T. Ha et al. 1997. A sensitive ELISPOT assay to detect low-frequency human T lymphocytes. *J. Immunol. Methods* 210:149-166.
6. Scheibenbogen, C., K.H. Lee, S. Mayer, S. Stevanovic, U. Moebius, W. Herr, H.G. Rammensee, and U. Keilholz. 1997. A sensitive ELISPOT assay for detection of CD8+ T lymphocytes specific for HLA class I-binding peptide epitopes derived from influenza proteins in the blood of healthy donors and melanoma patients. *Clin. Cancer Res.* 3:221-226.
7. Schmitt, A., U. Keilholz, and C. Scheibenbogen. 1997. Evaluation of the interferon-gamma ELISPOT-assay for quantification of peptide specific T lymphocytes from peripheral blood. *J. Immunol. Methods* 210:167-174.
8. Vaquerano, J.E., M. Peng, J.W. Chang, Y.M. Zhou, and S.P. Leong. 1998. Digital quantification of the enzyme-linked immunospot (ELISPOT). *BioTechniques* 25:830-835.
9. Yang, J., V.M. Lemas, I.W. Flinn, C. Krone, and R. Ambinder. 2000. Application of ELISPOT assay to the characterization of CD8+ responses to Epstein-Barr virus antigens. *Blood* 95:241-248.

This project was conducted within the Co-operative Research Centre for Vaccine Technology. Address correspondence to Dr. Mandvi Bharadwaj, Queensland Institute of Medical Research, The Bancroft Centre, 300 Herston Road, Brisbane 4029, Australia. e-mail: mandviB@qimr.edu.au

Received 8 May 2000; accepted 12 September 2000.

Mandvi Bharadwaj, Peter G. Parsons, and Denis J. Moss
*Queensland Institute of Medical Research
Brisbane, Australia*

Microplate Assay for the Measurement of Hydroxyproline in Acid-Hydrolyzed Tissue Samples

BioTechniques 30:38-42 (January 2001)

Collagens are major constituents of connective tissues. The organic matrix of bone, for example, consists of approximately 90% type I collagen (2) with a small amount of type III and other minor collagens. Typically, the fibrillar collagens contain approximately 100 hydroxyproline residues per 1000 amino acid residues (3), and measure-

Benchmarks

ment of this IMINO acid has been used to estimate total collagen content of tissues, preferably with consideration of the variability of their content of hydroxyproline.

Many methods have been developed for the measurement of hydroxyproline (4). Most commonly used methods rely on the oxidation of hydroxyproline to a pyrrole derivative, which reacts with *p*-dimethylaminobenzaldehyde (*p*-DAB) in an acid medium to give a colored product (the Ehrlich reaction). The color is enhanced by the presence of an organic solvent. Our simplified microplate assay is adapted from a segmented-flow auto-analysis procedure (5) based on a method originally described by Bergman and Loxley (1). The method described has been optimized for the measurement of hydroxyproline in acid hydrolysates of connective tissue, particularly mineralized bone powders, gelatin, and collagen tryptic digests. The assay was also found suitable for hydroxyproline determinations in nonmineralized connective tissues, including articular cartilage, meniscus, ligament, and loose connective tissue. In its present form, the assay is not suitable for the analysis of hydrolyzed urine samples.

We found that color development reached a plateau after 35 min at 60°C (Figure 1). We therefore decided to use an incubation period of 45 min, making the assay insensitive to uneven heating of the plate. Nonspecific color production was corrected by running blanks

without chloramine T and subtracting the average of the blank absorbances from the replicate means of the unknown samples and calibrators. We found that optical density values of the blanks were consistently small (8% of top calibrator) in tissue hydrolysates (intraassay CV 5.8%). We routinely incorporate several bone hydrolysate blanks into each assay plate. It is advisable to measure blanks for each different tissue or biological fluid to be analyzed.

A 96-well microplate (Appleton-woods, Birmingham, UK) was found to be suitable. This has individual wells joined with a web of plastic at the top of the plate rather than at the bottom, allowing water to circulate around each well. Small vents were cut into the sides of the plate to allow trapped air to escape from beneath the plate when being heated in the water bath.

Oxidation buffer contained of 600 mL isopropyl alcohol, 330 mL distilled water and 390 mL citrate/acetate buffer (containing 34 g sodium hydroxide, 34 g citric acid monohydrate, and 120 g sodium acetate trihydrate, adjusted to pH 6.0 with glacial acetic acid and made up to a final volume of 1 L). The oxidizing solution consisted of 300 mg chloramine T freshly dissolved in 50 mL oxidation buffer and kept in a dark reservoir. Ehrlich's reagent contained 6 g *p*-DAB (Park Scientific, Northampton, UK) dissolved in a mixture of 52 mL isopropyl alcohol and 16 mL 50% perchloric acid. Stock standard solutions of 1 mg/mL hydroxyproline in 10 mM HCl

were stored frozen in small portions and thawed as required to make working calibrator dilutions of 1, 2, 4, 6, 8, 10, and 12 µg/mL in 10 mM HCl. The working calibrators kept at 4°C were found to be stable for periods of up to one month. All reagents were obtained from BDH (Poole, UK) and were of general purpose reagent (GPR) grade unless otherwise stated.

Approximately 10 mg powdered bone tissue was hydrolyzed in 500 µL HCl (specific gravity 1.16, approximately 10 M) and 500 µL distilled water in screw-capped 6-mL polypropylene tubes (Sarstedt, Leicester, UK) at 108°C for 16 h. On cooling, the hydrolysates were partially neutralized with 4 mL sodium hydroxide (1 M) and centrifuged to sediment particulate debris. To monitor for both loss of hydroxyproline, due to over-hydrolysis, and the extent of hydrolysis, two internal standards were included in every batch. Each contained a replicate of a quality-control bone powder sample to which was added 500 µL either hydroxyproline (1 mg/mL) or gelatin (5 mg/mL).

All tissue hydrolysates were diluted 1:20 before assay. Calibrators, unknown samples, internal standards, and blanks were pipetted, in triplicate, in 50-µL aliquots onto the 96-well plate. The calibrators were positioned in ascending and descending order at both ends, and a third was set diagonally across the plate. All the other samples were pipetted in random positions on the plate. An oxidizing solution con-

Figure 1. Development of optical density with time of incubation at 60°C for the 2 µg/mL (filled circles) and 12 µg/mL (filled triangles) calibrators.

Figure 2. Typical calibration curve with each point representing individual measurements. The mean values were used to fit the least-squares correlation line.

Benchmarks

taining chloramine T (100 μ L) was added to each well except those designated as blanks. The same amount of oxidation buffer without chloramine T was added to each blank well. The plate was mixed by gentle shaking and left for 5 min. Ehrlich's reagent (100 μ L) was added to each well and mixed thoroughly. The plate was covered with an adhesive plate seal and incubated in a water bath at 60°C for 45 min. Absorbance was read at 570 nm using an MRX microplate reader (Dynatech Laboratories, West Sussex, UK). The average optical densities (OD) for the calibrators and unknown samples were corrected for the average blank OD by subtraction. A calibration curve (Figure 2) was constructed, and the hydroxyproline concentration of the unknown samples was calculated from the "best-fit" equation of the line.

The within-run CV was 4.3%, 3.8%, and 4.1% for the concentration ranges, 2–4.5, 5–8, and 8–10 μ g/mL hydroxyproline, respectively. The interassay CV in the lower range was 9.3% and was calculated from 16 runs with repeated hydrolysis. The sensitivity of the assay was determined as the concentration at which the standard deviation of the result was half of that result. Using this criterion, the sensitivity of the assay is 0.1 μ g/mL hydroxyproline, and it is linear up to 40 μ g/mL. The recovery of added hydroxyproline was $100 \pm 8\%$ (range 88%–116%), and the recovery of added gelatin was $93 \pm 12\%$ (range 75%–115%) based on a nominal hydroxyproline content.

This assay for hydroxyproline in connective tissue hydrolysates shows an improved reproducibility compared with the automated technique described by Worsfold et al. (5), which has intra- and interassay CVs of 5.9% and 12%, respectively (unpublished data), although those statistics were compiled over several years of urine assays that have variable blank optical densities. The procedure is convenient and allows increased sample throughput using minimum sample and reagent volumes. The equipment used is nowadays more widely available than segmented flow analyzers. The method as described did not work well for urines, which contain less hydroxyproline and would not be diluted after hydrolysis, so that larger amounts of hy-

drochloric acid and sodium chloride are carried forward into the assay.

REFERENCES

1. **Bergman, I. and R. Loxley.** 1963. Two improved and simplified methods for the spectrophotometric determination of hydroxyproline. *Anal. Chem.* 35:1961-1965.
2. **Bouvier, M.** 1989. The biology and composition of bone, p. 1-13. *In* S.C. Cowin (Ed.), *Bone Mechanics*. CRC Press, Boca Raton, FL.
3. **Dickson, I.R. and M.K. Bagga.** 1985. Changes with age in the non-collagenous proteins of human bone. *Connect. Tissue Res.* 14:77-85.
4. **Mitchell, A.D. and I.E.P. Taylor.** 1970. The spectrophotometric determination of hydroxyproline: an analytical investigation. *Analyst* 95:1003-1011.
5. **Worsfold, M., M.W.J. Davie, and M.J. Hadaway.** 1999. Age-related changes in body composition, hydroxyproline, and creatinine excretion in normal women. *Calcif. Tissue Int.* 64:40-44.

Address correspondence to Ms. Sharon Brown, Charles Salt Centre, Robert Jones & Agnes Hunt Orthopaedic Hospital, Oswestry, Shropshire SY10 7AG, UK. e-mail: sharon.rowbotham@rjahoh-tr.wmids.nhs.uk

Received 30 May 2000; accepted 25 September 2000.

Sharon Brown^{1,2}, Michael Worsfold¹, and Christopher Sharp¹

¹*Robert Jones & Agnes Hunt Orthopaedic Hospital*

Oswestry, Shropshire

²*University College Chester Chester, Cheshire, UK*

Use of DAPI Staining for Quantitation of Cell Chemotaxis

BioTechniques 30:42-48 (January 2001)

Cell migration is a fundamental biological phenomenon that plays a pivotal role in developmental differentiation, inflammatory responses, wound heal-

ing, and tumor metastasis. Cells migrate in response to concentration gradients of diverse protein growth factors such as platelet-derived growth factor (PDGF) and vascular endothelial growth factor, as well as lipid mediators such as sphingosine 1-phosphate (SPP), lysophosphatidic acid, and arachidonic acid. The ability of cells to react chemotactically to stimuli is of particular importance in angiogenesis, the process whereby new blood vessels are formed (2).

A variety of methods exist by which the migratory behavior of cells can be evaluated. These methods include labeling of live cells with fluorescent and luminescent tags (5,13), the use of cell type-specific stains (11), or the use of broad specificity stains such as hematoxylin-eosin mixture (9). Some of these methods are complex and require specialized genetic constructions or equipment, while others are time consuming and cumbersome.

Our current investigations have focused on the characterization of chemotactic responses of endothelial cells and smooth muscle cells to clarify the essential aspects of angiogenic responses. These two cell types are intimately involved in formation of capillaries and mature vessels (1,3,6,7). The migration of these cells toward their specific chemoattractants, SPP and PDGF, respectively, is routinely measured by counting cells that traverse the filter separating the upper compartment of a chemotactic chamber from the bottom well containing the stimulus. The cells that reach the lower surface of the filter are usually stained with hematoxylin and eosin and enumerated microscopically.

Although the hematoxylin-eosin staining is technically simple, the process is rather cumbersome and requires many steps to complete (Table 1). More importantly, when a vigorous chemotactic response is quantified, the cells of interest are not easy to enumerate accurately because of difficulties involved in determining the boundaries between closely adjacent cells and because of the irregular shape of the cells. We have developed a simple and rapid method to overcome these difficulties by employing a 4', 6-diamidino-2-phenylindole (DAPI) (Sigma, St. Louis, MO, USA) staining technique. DAPI intercalates