Benchmarks

- their association with detergent-resistant membranes. *J. Cell Biol.* 145:877-887.
12. **Smallcombe, A.** 2001. Multicolor imaging: the important question of co-localization. *BioTechniques* 30:1240-1246.
13. **Stauffer, T.P. and T. Meyer.** 1997. Compartmentalized IgE receptor-mediated signal transduction in living cells. *J. Cell Biol.* 139:1447-1454.

This work was supported by American Heart Association Beginning Grant-in-Aid no. 9960341Z. Address correspondence to Dr. William Rodgers, Molecular Immunogenetics Program, Oklahoma Medical Research Foundation, 825 NE 13th St. MS 17, Oklahoma City, OK 73104, USA. e-mail: william-rodgers@omrf.ouhsc.edu

Received 6 June 2001; accepted 9 October 2001.

William Rodgers
*Oklahoma Medical Research
Foundation
Oklahoma City, OK, USA*

**For reprints of this or
any other article, contact
Reprints@BioTechniques.com**

Fluorescence Polarization Standards for High-Throughput Screening and Imaging

BioTechniques 32:34-42 (January 2002)

Fluorescence polarization has been utilized for some time in the form of fluorescence polarization immunoassays for therapeutic drugs, drugs of abuse, and other small molecules (3, 11), and more recently for immunoassays for macromolecules (9). Polarization-based assays have been developed for high-throughput screening applications, particularly lead identification. These assays include protease assays (24), kinase assays (2,22,23), and DNA hybridization reactions (20,21,27), as

well as assays for free metal ions (26) and glucose (16). These assays exploit the unique advantages of fluorescence polarization-based assays in general: good sensitivity, rapid response, homogenous format (no separation required), simple instrumentation, and a robust format that minimizes artifacts. The robust nature of the polarization measurement arises from its ratiometric character, readily achieved in the classic T-format of Weber (28). The polarization of fluorescence can be expressed in terms of the anisotropy (r) or the polarization (p):

$$p = (I_{\parallel}/I_{\perp} - 1)/(I_{\parallel}/I_{\perp} + 1) \quad [\text{Eq. 1.1}]$$

$$r = (I_{\parallel}/I_{\perp} - 1)/(I_{\parallel}/I_{\perp} + 2) \quad [\text{Eq. 1.2}]$$

where $I_{\parallel}$ is the intensity of the fluorescence emission with the emission polarizer oriented parallel to the vertically oriented excitation polarizer, and $I_{\perp}$ the intensity with the emission polarizer oriented perpendicular (horizontal) to the excitation. Like other ratiometric fluorescence measurements, polarization measurements can thus be insensitive to fluctuations in excitation intensity, fluorophore concentration, sample thickness, and some inner filter effects. Therefore, it is well suited for quantitative assays and also imaging applications such as fluorescence polarization microscopy (1,5,6). Note that polarization measurements taken in the more common L-format are not strictly ratiometric since the intensities in Equation 1 are measured serially; however, most excitation sources vary only a few tenths of 1% over a few minutes, so the precision can be quite good. Although we will refer mainly to fluorescence polarization as opposed to fluorescence anisotropy since the former has come into wide use in the high-throughput screening community, it is nevertheless true that, although the terms are synonymous and algebraically easy to interconvert, anisotropy is much more tractable theoretically (see Equation 2).

The need for and utility of standards in all forms of chemical analysis are widely accepted, and the same is true for fluorescence polarization assays. The availability of convenient standards permits calibration of the response of the instrument in terms of the analyte amount or concentration; veri-

fies the correct functioning of the instrument, particularly following alignment or repair; and provides a basis for troubleshooting anomalous results. For some clinical assays, frequent recalibration is required by law. Ideally, a standard material should be stable, easy to formulate, transportable, readily transferable among laboratories, traceable to some primary standard, and inexpensive. The standard should be usable under conditions that closely mimic (if they are not identical to) the measurement of "real" samples.

Although fluorescence polarization measurements are easy to make and relatively free from artifact, some agents and influences may perturb a polarization measurement, especially in 96-well plates and similar formats. In general, the extent of fluorescence polarization from a particular sample depends primarily on the ratio of the fluorescence lifetime to the rotational correlation time, according to the Perrin equation:

$$(1/p - 1/3) = (1/p_0 - 1/3)(1 + \tau/\theta_c) \quad [\text{Eq. 2.1}]$$

$$r/r_0 = 1/(1 + \tau/\theta_c) \quad [\text{Eq. 2.2}]$$

where τ is the lifetime, θ_c is the rotational correlation time, p is the measured polarization, and p_0 is the polarization measured in the absence of rotational diffusion, ordinarily 0.5. Among these influences is temperature, which can affect both rotational diffusion and fluorescence lifetime. However, temperature can readily be controlled. Similarly, the viscosity of the solvent can be controlled by its composition and temperature. Also of potential concern is scattered light (Rayleigh or Raman), which, because of its high polarization (1.0), can readily corrupt measurements even when it represents only a small fraction of the light hitting the detector. Again, this problem can be avoided by careful optical filtration of the emission. Another potential source of error is collisional quenching of the fluorophore by various solutes because it affects the lifetime. However, unless the fluorescence lifetime is long, it will be little affected by this phenomenon unless the quenching solutes are present at nearly millimolar concentrations, or in certain rare cases. Again, avoidance of such solutes at high levels minimizes the problem. The perturba-

tion to measured polarization attributable to the typical inability of the emission optical train and detector to transmit and detect the two orthogonal polarizations equally well is less well known than the *g*-factor universally used to correct for it. The issue of polarizer nonideality and misalignment is subtler, where two polarizers in orthogonal orientations or one being rotated are mismatched or misaligned so that they do not pass the two polarizations equally well. On the emission side, this can be corrected by *g*-factor, but on the excitation side is more troublesome. The procedure for polarizer alignment is described by Lakowicz (15). In addition, the presence of fluorescence backgrounds in tissues and other specimens corrupts measurements, but this background can often be subtracted or the background avoided altogether by the use of long wavelength dyes (25). Fi-

nally, certain optical components may be birefringent and thereby corrupt measurements (among these are many transparent plastics, microscope objectives, and optical fiber) under conditions of even mild stress. The depolarization of fluorescence caused by energy transfer among like molecules at high concentrations (homotransfer) has been known for many years but is unlikely to be important except at high micromolar concentrations or in special circumstances (7). Of course, there is the well-known excitation wavelength dependence of fluorescence polarization, which is discussed below. For more detailed discussion of these issues, see Lakowicz (15).

Previous standards for polarization measurements have been less than satisfactory. Many workers have used Rayleigh scatterers as standards because they are very convenient, and the

polarization of (once-scattered) light is 1.000. However, this may be harder to arrange in the front-face geometry of many plate readers because the Fresnel reflections off other surfaces may be admixed with the Rayleigh scatter from the sample. Also, care must be taken to use minimum levels of scatterer to avoid multiple scattering events, which depolarize emission. The necessary amount varies dramatically with excitation wavelength because scattering increases as the fourth power of the light frequency. At the long wavelengths of probes such as CY-5.5, the amount of scatterer necessary increases correspondingly. Typical scattering solutes include colloidal silica, ordinary coffee creamer, and plastic beads. Some have also used Raman scattering as a reference, which avoids the problem of double scattering events. However, Raman is a relatively weak emission, and other

Figure 1. Absorbance and emission spectra of standards. Absorbance (dashed lines) and fluorescence emission spectra (solid lines) of Rose Bengal (upper left), Phloxine B (upper right), Erythrosin B (lower left), and Ru(bipy)₃ (lower right) are depicted.

Benchmarks

noise sources and background fluorescence can corrupt these measurements.

Some fluorophores have also been used as polarization standards. Some investigators have used fluorescein in aqueous solution, which, because of its moderate lifetime of about 4 ns and short rotational correlation time of perhaps 100 ps exhibits a polarization of 0.023 (10). This value is not well established, however, which would make it less useful as a standard. In particular, inspection of Equation 1 shows that for this polarization the difference in intensities between $I_{\parallel}$ and $I_{\perp}$ is less than 2% (in an L-format instrument that measures the intensities serially, rather than as a ratio), which is only a few fold better than the ordinary precision of a fluorescence intensity measurement. A better standard for a low-polarization sample might be the same fluorescein sample but excited at a wavelength about 400 nm where its p_0 is close to zero (10); therefore, the average value of the difference in intensities should be zero. Still better would be a fluorophore with a very long lifetime in solution because its anisotropy would be very close to zero over a broad range of excitation, and it might therefore be easier to match conditions with the experiment of interest. Such a compound is ruthenium tris(bipyridyl) [$\text{Ru}(\text{bipy})_3$], which has a very long-lived metal-to-ligand charge transfer excited state (4), resulting in effectively a zero polarization almost throughout its excitation band.

Perhaps a better standard would be a fluorophore with a short enough lifetime that even in fluid solvents it would exhibit a significant polarization. The archetypes of these compounds are the stilbenes and styrenes introduced by Kowski and colleagues, which have lifetimes of as little as a few picoseconds and commensurately high polarizations in fluid solvents (8,12,13,17). Unfortunately, these compounds do not excite well at wavelengths beyond 500 nm, they are largely insoluble in water, and their quantum yields are low, making them difficult to use in aqueous solvents because of the high concentrations required. Also, they are not commercially available at present.

However, three other compounds (Phloxine B, Rose Bengal, and Ery-

throsin B) are more attractive in this regard, and we believe they have the potential to be useful standards. All are halogenated derivatives of fluorescein that have good water solubility and high absorbance in the visible; their spectra are depicted in Figure 1. The "heavy atom effect" of the bromine atoms in Phloxine B and the iodine atoms in Erythrosin B and Rose Bengal is to promote efficient spin-orbit coupling and consequently induce a spin flip into the triplet manifold, effectively quenching the fluorescence. The result is much shorter lifetimes [for Erythrosin B in water, 75 ps (19), and for Rose Bengal in water, 76 ps (18)] and consequent substantial polarization even in aqueous solution. The short lifetimes make collisional quenching negligible even in the presence of known quenchers. Erythrosin B, Rose Bengal, and Phloxine B are available at a certified level of purity for use as stains and are quite inexpensive. While stable dry at room temperature, all are photosensitizers and are unstable in dilute solutions exposed to light. Stock solutions at higher concentrations stored in the dark are usable for long periods. Like fluorescein solutions that must be dilute ($A < 0.05$) to avoid self-quenching and homotransfer effects that would corrupt polarization measurements, working solutions of these dyes should also have absorbances less than 0.2, although they are less susceptible to such effects because of their short lifetimes [except $\text{Ru}(\text{bipy})_3$, for which it is a non-issue].

Phloxine B (Aldrich cat. no. 19827-7; Acid red 92; CI 45410) Rose Bengal (Aldrich 19,825-0; Acid red 94; CI 45440) Acid Erythrosin B, (Aldrich 19,826-9; Acid red 51; CI 45430), and $\text{Ru}(\text{bipy})_3$ (Aldrich 22,475-8) (Figure 1) were conveniently dissolved in ethanol or water to approximately 1 mg/mL, and aliquots of these solutions were then diluted in 50 mM HEPES, pH 7.2, to absorbances of 0.2 or less at the peak of absorbance. Polarizations were measured in L-format on a ThermoSpectronics AB-2 at 25°C using film polarizers or an SLM 8000 with a double excitation monochromator and Glan Thompson calcite polarizers under the spectral conditions listed in Figure 2, and each is the average of 10 measure-

ments. The excitation polarization spectra depicted in Figure 2 indicate the wavelength ranges of excitation of each of the fluorophores wherein the indicated values of the polarization may be measured. The resulting values from the two instruments agreed to within the standard deviation of the measurements (ordinarily ± 0.002). Blank spectra of the buffer showed no fluorescence under these conditions. Blank spectra are strongly recommended inasmuch as the fluorescence quantum yields of the fluorophores are rather low and the ability of background fluorescence to contribute significantly even at low levels is of concern.

Fluorescence polarizations were measured for the samples of Phloxine B, Rose Bengal, $\text{Ru}(\text{bipy})_3$, and Erythrosin B at 25°C under the spectral conditions described in Figure 2. The data in Figure 2 are depicted as excitation polarization spectra to indicate where in the excitation range the maximum r_0 may be measured. The standard deviations of the individual intensity measurements used to calculate the polarization/anisotropy values in Figure 2 typically ranged from 0.4% to 0.7%. Workers may not observe exact correspondence with these values because of slightly different spectral conditions in their instruments or viscosities of buffers used.

Several factors may cause the measured values for the standard(s) to differ significantly from values given here, and some of these factors depend on the design of particular instruments. Certainly the standards should be measured initially when the instrument is known to be operating properly and well aligned, for instance, when new or after servicing. If the precision of values as judged by the standard deviation is much worse than specified or observed historically, then potential causes include scattering from incompletely dissolved standards or dust, electronic noise from detectors or amplifiers, or noise from an aging excitation source. Xenon lamp sources should vary less than 0.5% over several minutes time; more noise or any noise spike is grounds for replacement. Photomultiplier tube (PMT) detectors should be stable (under steady illumination) at low frequencies to within 1%; note that

Benchmarks

exposure of PMTs to excessive light under high voltage can increase their noise, if it does not destroy them altogether. LED sources, photodiodes, and CCD cameras are typically very stable. If the values are inaccurate, the most likely cause is misalignment of the excitation and/or emission optical train. First check for looseness or obvious displacement in mounting of optical components or the sample. Changes in the g-factor are independent of excitation and consequently would indicate a change in the emission optical train, including the detector. If the polarizers are designed to rotate in the instrument, their alignment (around the optical axis)

can be checked according to Lakowicz (14). Note that displacement of the sample such that the path of excitation or emission changes can change the apparent polarization. A special case of this is caused by using too high a fluorophore concentration, which causes an inner filter effect whereby most excitation is absorbed at the surface, which may be suboptimal for collection of light by the emission optical train. If the measured polarizations are much higher than shown (particularly if they are greater than 0.5), the excitation and emission filtration should be checked to assure that Rayleigh or Raman scattered light is not reaching the detector.

Having standards that exhibit polarizations ranging from 0.000 to 0.335 is useful. First, the higher value is useful because any artifacts that tend to depolarize the emission will reduce this number, whereas the change might not be apparent for a lower polarization. Similarly, the zero value makes it easy to detect if the g-factor is being measured correctly because the values of $I_{||}$ and $I_{\perp}$ must be equal. The compounds are cheap enough to be included in micro-well plates as standards as often as desired. The compounds are widely available, not notably hazardous, and easily formulated. Spectroscopically, they closely mimic [except for Ru(bipy)₃]

Figure 2. Excitation polarization spectra of standards. Excitation spectra of Rose Bengal (upper left), Phloxine B (upper right), Erythrosin B (lower left), and Ru(bipy)₃ (lower right) are depicted. In each panel, the recommended excitation wavelength range is indicated where the polarization is largely independent of wavelength and the accurate values indicated in each panel may be obtained.

Benchmarks

the substituted Rhodamine and cyanine dyes now in wide use. They are compatible with aqueous buffers in dilute solution, so the standard can usually be in the same buffer as the cells or samples whose polarization is being measured. Finally, the availability of convenient standards should promote the use of higher-precision instruments, such that systems exhibiting only modest polarization changes can still be used for quantitative analyses.

REFERENCES

1. Axelrod, D. 1989. Fluorescence polarization microscopy, p. 333-352. *In* D.L. Taylor and Y.-L. Wang (Eds.), *Methods in Cell Biology*, vol. 30. Academic Press, New York.
2. Coffin, J., M. Latev, X. Bi, and T.T. Nikiforov. 2000. Detection of phosphopeptides by fluorescence polarization in the presence of cationic polyamino acids: application to kinase assays. *Anal. Biochem.* 278:206-212.
3. Dandliker, W.B., R.J. Kelly, J. Dandliker, J. Farquhar, and J. Levin. 1973. Fluorescence polarization immunoassay. Theory and experimental method. *Immunochemistry* 10:219-227.
4. Demas, J.N. and G.A. Crosby. 1971. Quantum efficiencies of transition metal complexes. II. Charge-transfer luminescence. *J. Am. Chem. Soc.* 93:2841.
5. Dix, J.A. and A.S. Verkman. 1990. Mapping of fluorescence anisotropy in living cells by ratio imaging: application to cytoplasmic viscosity. *Biophys. J.* 57:231-240.
6. Fushimi, K., J.A. Dix, and A.S. Verkman. 1990. Cell membrane fluidity in the intact kidney proximal tubule measured by orientation-independent fluorescence anisotropy imaging. *Biophys. J.* 57:241-254.
7. Gaviola, E. and P. Pringsheim. 1924. [First observation of homotransfer]. *Zeitschrift fur Physik.* 24:24.
8. Gloyna, D., A. Kowski, I. Gryczynski, and H. Cherek. 1987. The influence of solvent polarity and viscosity on the deactivation of the S1-state of donor-acceptor substituted trans-stilbenes. *Monatshefte fur Chemie.* 118:759-772.
9. Guo, X.-Q., F.N. Castellano, L. Li, and J.R. Lakowicz. 1998. Use of a long-lifetime Re(I) complex in fluorescence polarization immunoassays of high-molecular-weight analytes. *Anal. Chem.* 70:632-637.
10. Jameson, D.M., G. Weber, R.D. Spencer, and G. Mitchell. 1978. Fluorescence polarization: measurements with a photon-counting photometer. *Rev. Sci. Instrum.* 49:510-514.
11. Jolley, M.E., S.D. Stroupe, K.S. Schwenzer, C.J. Wang, M. Lu-Steffes, H.D. Hill, S.R. Popelka, J.T. Holen, and D.M. Kelso. 1981. Fluorescence polarization immunoassay III: an automated system for drug determination. *Clin. Chem.* 27:1575-1579.
12. Kowski, A., I. Gryczynski, and Z. Gryczynski. 1993. Fluorescence anisotropies of 4-dimethylamino-w-diphenylphosphinyl-trans-styrene in isotropic media in the case of one- and two-photon excitation. *Zeitschrift fur Naturforschung.* 48a:551-556.
13. Kowski, A., I. Gryczynski, K. Nowaczyk, P. Bojarski, and J. Lichacz. 1991. Deactivation of the S1-state of w-substituted 4-dimethylamino-trans-styrenes in alkane solutions. *Zeitschrift fur Naturforschung.* 46a:1043-1048.
14. Lakowicz, J.R. 1983. *Principles of Fluorescence Spectroscopy*. Plenum Press, New York.
15. Lakowicz, J.R. 1999. *Principles of Fluorescence Spectroscopy*, 2nd ed. Kluwer Academic/Plenum Publishers, New York.
16. Lakowicz, J.R., I. Gryczynski, Z. Gryczynski, L. Tolosa, L. Randers-Eichhorn, and G. Rao. 1999. Polarization-based sensing of glucose using an oriented reference film. *J. Biomed. Optics* 4:443-449.
17. Lakowicz, J.R., I. Gryczynski, G. Laczo, and D. Gloyna. 1991. Picosecond fluorescence lifetime standards for frequency- and time-domain fluorescence. *J. Fluorescence* 1:87.
18. Lakowicz, J.R., G. Laczo, and I. Gryczynski. 1986. 2-GHz frequency-domain fluorometer. *Rev. Sci. Instrum.* 57:2499.
19. Matczuk, A., P. Bojarski, I. Gryczynski, J. Kusba, L. Kulak, and C. Bojarski. 1995. The influence of water structure on the rotational depolarization of fluorescence. *J. Photochem. Photobiol. A Chem.* 90:91-94.
20. Murakami, A., M. Nakaura, Y. Nakatsuji, S. Nagahara, Q. T. Cong, and K. Makino. 1991. Fluorescent-labeled oligonucleotide probes: detection of hybrid formation in solution by fluorescence polarization spectroscopy. *Nucleic Acids Res.* 19:4097-4102.
21. Nikiforov, T.T. and S. Jeong. 1999. Detection of hybrid formation between peptide nucleic acids and DNA by fluorescence polarization in the presence of polylysine. *Anal. Biochem.* 275:248-253.
22. Seethala, R. and R. Menzel. 1998. A fluorescence polarization competition immunoassay for tyrosine kinases. *Anal. Biochem.* 255:257-262.
23. Seethala, R. and R. Menzel. 1997. A homogeneous, fluorescence polarization assay for src-family tyrosine kinases. *Anal. Biochem.* 253:210-218.
24. Spencer, R.D., F.B. Toledo, B.T. Williams, and N.L. Yoss. 1973. Fluorescence polarization assay of protease. *Clin. Chem.* 19:838-844.
25. Thompson, R.B. 1994. Red and near-infrared fluorometry, p. 151-181. *In* J.R. Lakowicz (Ed.), *Topics in Fluorescence Spectroscopy: Probe Design and Chemical Sensing*, vol. 4. Plenum Press, New York.
26. Thompson, R.B., B.P. Maliwal, and C.A. Fierke. 1998. Expanded dynamic range of free zinc ion determination by fluorescence anisotropy. *Anal. Chem.* 70:1749-1754.
27. Walker, G.T., J.G. Nadeau, C.P. Linn, R.F. Devlin, and W.B. Dandliker. 1996. Strand displacement amplification (SDA) and transient-state fluorescence polarization detection of *Mycobacterium tuberculosis* DNA. *Clin. Chem.* 42:9-13.
28. Weber, G. 1956. Photoelectric method for the measurement of the polarization of the fluorescence of solutions. *J. Opt. Soc. Am.* 46:962.

The authors thank the National Science Foundation, Office of Naval Research, and National Institutes of Health for support, Professor Joseph Lakowicz for his continuing encouragement, and Dr. Susan Keating of Genentech for helpful comments on the manuscript. Address correspondence to Dr. Richard B. Thompson, Department of Biochemistry and Molecular Biology and Center for Fluorescence Spectroscopy, University of Maryland School of Medicine, 108 N. Greene Street, Baltimore, MD 21201, USA. e-mail: rthomps@umaryland.edu

Received 3 August 2001; accepted 19 October 2001.

**Richard B. Thompson,
Ignacy Gryczynski,
and Joanna Malicka**
*University of Maryland
School of Medicine
Baltimore, MD, USA*