Benchmarks

gests the bacterially methylated template plasmid DNA at 5'-GATC-3' sequences and leaves newly synthesized, unmethylated DNA intact. Two microliters of the reaction are electroporated into XL-1 Blue *E. coli* (Stratagene, La Jolla, CA, USA), and transformants are selected on LB agar plates containing the appropriate antibiotic. Plasmids from the colonies are then sequenced using standard techniques.

This method is highly versatile and not only allows transfer of single-point mutations but also multiple mutations simultaneously if the original mutation donor plasmid has more than one sequence change from the recipient. In addition, we have also used two site-specific mutation-containing primers to incorporate two mutations simultaneously into the ends of the first-round PCR product. Both mutations can then be transferred to the recipient plasmid in the second round of PCR. Again, this allows incorporation of two (or possibly more) mutations into the final vector simultaneously rather than sequentially. This is especially useful if the distance between the two desired mutations is greater than can be handled effectively with synthetic oligonucleotides. In addition to point mutation transfer, it seems feasible that this method might also prove useful in generating chimeric DNAs for domain swapping or even whole gene swapping, as long as the two DNA regions to be swapped are of a similar enough sequence and size. For example, it may be possible to swap the small (approximately 400 bp) E6 and E7 viral oncogenes between two strains of papillomavirus since this virus contains double-stranded, circular DNA approximately 8 kb in length.

One limitation of this technique that we have seen is the length of the initial PCR product. For example, a product of 300 bp worked well, while 600-bp products worked poorly, if at all. However, this is only a minor limitation for a mutation transfer procedure since 100 amino acids is a fairly large region for mutational investigation. Furthermore, if the mutations are more widely dispersed, several additional primer sets can be designed to more fully encompass the overall regions to be transferred. Overall, we believe this procedure is a useful adjunct to other mutation transfer

procedures and will often prove simpler and less expensive than subcloning or traditional site-specific mutagenesis.

REFERENCES

1. **Chen, G. and A. Stenlund.** 2000. Two patches of amino acids on the E2 DNA binding domain define the surface for interaction with E1. *J. Virol.* 74:1506-1512.
2. **Gonzalez, A., C. Bazaldua-Hernandez, M. West, K. Woytek, and V.G. Wilson.** 2000. Identification of a short, hydrophilic amino acid sequence critical for origin recognition by the bovine papillomavirus E1 protein. *J. Virol.* 74:245-253.
3. **Leng, X., J.H. Ludes-Meyers, and V.G. Wilson.** 1997. Isolation of an amino-terminal region of bovine papillomavirus type 1 E1 protein that retains origin binding and E2 interaction capacity. *J. Virol.* 71:848-852.
4. **Sarafi, T.R. and A.A. McBride.** 1995. Domains of the BPV-1 E1 replication protein required for origin-specific DNA binding and interaction with the E2 transactivator. *Virology* 211:385-396.
5. **Spear, M.** 2000. Efficient DNA subcloning through selective restriction endonuclease digestion. *BioTechniques* 28:660-668.
6. **Thorner, L.K., D.A. Lim, and M.R. Botchan.** 1993. DNA-binding domain of bovine papillomavirus type 1 E1 helicase: structural and functional aspects. *J. Virol.* 67:6000-6014.
7. **Tripodi, M., S. Perfumo, R. Ali, L. Amicone, C. Abbott, and R. Cortese.** 1990. Generation of small mutation in large genomic fragments by homologous recombination: description of the technique and examples of its use. *Nucleic Acids Res.* 18:6247-6251.
8. **Watkins, B.A., A.E. Davis, F. Cocchi, and M.S. Reitz, Jr.** 1993. A rapid method for site-specific mutagenesis using larger plasmids as templates. *BioTechniques* 15:700-704.

We wish to thank Julian Leibowitz and Rajesh Miranda for helpful discussions during the development of this work. This work was funded through American Cancer Society grant no. VM 183. Address correspondence to Dr. Van G. Wilson, Department of Medical Microbiology and Immunology, Texas A&M University System Health Science Center, College Station, TX 77843-1114, USA. e-mail: v-wilson@tamu.edu

Received 17 July 2001; accepted 27 September 2001.

**Michael West and
Van G. Wilson**

*Texas A&M University System
Health Science Center
College Station, TX, USA*

Direct Fluorescent Primers Are Superior to M13-Tailed Primers for *Pinus taeda* Microsatellites

BioTechniques 32:46-52 (January 2002)

Microsatellite marker data can be automated for high-throughput mapping projects, and automated collection methods are superior to PAGE with silver staining or radiolabeling in time, resolution, and safety. Many microsatellites can be loaded in a single lane using multiple dyes with different wavelength emissions and product size classes. The sensitivity of the automated system to fluorescently labeled PCR products reduces the amount of the required product. Automated systems have high resolution, detecting a single base-pair difference. Fluorescent dyes used in automated systems are safer and more stable than silver-staining protocols or radioisotopes.

Pine microsatellites can be especially difficult for developing high-throughput protocols because *Pinus* spp. genomes range from 18600 to 30000 Mb per haploid genome (8) and, similar to many large plant genomes, are composed mostly of highly repetitive DNA (76%–86%) (5,10). Gene families are also highly duplicated (9), and duplication includes families of clustered microsatellites (7). Family members have conserved primer-binding sites and even conserved flanking regions, but repeat composition is highly variable within a family. Mispriming or suboptimal PCR amplification can result in no amplification, random priming, or amplification of non-orthologous members of the same microsatellite family. The sources of PCR-banding artifacts are highly duplicated gene families, high-copy microsatellite families, and highly repetitive DNA.

We compared the M13-tailed primer method and the direct fluorescent primer method for high-throughput microsatellite data collection using the IR² Gene Read IRTM 4200 DNA Analyzer (LI-COR, Lincoln, NE, USA). The M13-tailed primer method is based on a

Benchmarks

forward primer for each microsatellite primer set designed with a 5'-tail. The 5'-tail has the M13 sequence 5'-CAC-GACGTTGTAAAACGAC-3' (11,14). A third M13 primer carries the fluorescent label that binds to the forward primer during the PCR amplification. The M13-primer method circumvents the need for a fluorescently labeled primer for each microsatellite primer set. In mammalian species, the M13-tailed primer method is superior to direct fluorescent primer labeling (1). The direct fluorescent primer method uses a fluorescently labeled forward primer for each microsatellite primer set.

The *Pinus taeda* microsatellites were previously reported to have single PCR products, stable inheritance, and multiple alleles per locus (6,16). We tested a multiple-allele pedigree using a three-generation outbred pedigree *P. taeda* (2) with four unrelated grandparents, two parents, and 118 offspring. Genomic DNA was extracted from fresh embryo tissue and lyophilized needle tissue (3). One hundred polymorphic nuclear microsatellites were randomly selected from a list of *P. taeda* microsatellites (<http://forestry.tamu.edu/genetics>). Of these, 80 could be amplified with the M13-tailed primer method.

We used standard PAGE as a control. Primer amplification and PCR product size were first tested with PAGE gels without fluorescent dye labeling. Polymorphism and Mendelian inheritance were determined using four grandparents and two parents of same outbred pedigree (2). The mating type configuration was classified for each locus (15). Microsatellites were classified as follows: 24 were in backcross configuration for two alleles (BC), 11 were in backcross configuration for three alleles (MBC), 9 were in intercross configuration at two alleles (IC), and 36 were in intercross configuration at either three or four alleles (MIC). PCR products were separated on 6% polyacrylamide denaturing gels using modifications from Echt et al. (4). Gels were run at a constant temperature of 55°C for 1.5–3.5 h, followed by silver staining (4,5).

The M13-tailed primer method had a 19-bp sequence directly labeled with an infrared fluorophore, IRDye™700 or 800 (LI-COR). A reaction for the

M13-tailed primer method includes three primers: (i) the M13 forward primer (-29) to which the 19-bp M13-tailed sequence has been added, (ii) the unlabeled reverse primer, and (iii) the M13 infrared fluorescent dye-labeled primer. A primary PCR product is generated using the M13-tailed forward primer and the unlabeled reverse primer. The secondary PCR product is produced using the M13 infrared fluorescent dye-labeled primer and the unlabeled reverse primer. It is the IR-labeled PCR products generated from the secondary amplification that are detected by the DNA analyzer.

Re-annealing temperatures and MgCl₂ concentrations for the M13-tailed primer method were optimized before marker data collection. Optimum T_m using M13-tailed primers was tested using a gradient thermal cycler (RoboCycler® Gradient 96; Stratagene, La Jolla, CA, USA). MgCl₂ concentration for M13-tailed primers was optimized by testing 0.5, 1.0, 2.0, 2.5, 3.0, 3.5, and 4 mM. Amplifications were performed in a 10-μL reaction volume consisting of 10 ng template DNA, 50 mM Tris, pH 9.0, 20 mM (NH₄)₂SO₄, 100 μg/μL gelatin, 1.0% DMSO, 200 μM dNTPs, 1.0 pmol M13-tailed forward primer, 1.0 pmol reverse primer, 0.5 pmol IRDye700 or 800-labeled M13-tailed primer, 4.0 mM MgCl₂, and 0.25 U *Taq* DNA polymerase.

The direct fluorescent primer method was the other high-throughput method tested. For each microsatellite, the forward primer was synthesized and then conjugated to one of two infrared fluorescent dyes, IRDye700 or 800, at its 5' end (MWG-Biotech, Charlotte, NC, USA). PCR amplification using a labeled forward primer and an unlabeled reverse primer yielded fluorescently labeled PCR products.

The PCR protocol for the direct fluorescent primer method differed from the M13-tailed primer method by the use of a pair of primers: 2.0 pmol IRDye700 or 800-labeled forward primer and 2.0 pmol unlabeled reverse primer. The reactions were performed in a Tetrad™ Gradient Thermal Cycler (MJ Research, Waltham, MA, USA).

Amplifications were performed in a 10-μL reaction volume consisting of 10 ng template DNA, 50 mM Tris, pH 9.0,

20 mM (NH₄)₂SO₄, 100 μg/μL gelatin, 1.0% DMSO, 200 μM dNTPs, 1.0 pmol reverse primer, 1.0 pmol IRDye-labeled forward primer, 4.0 mM MgCl₂, and 0.25 U *Taq* DNA polymerase. The IRDye700 and 800-labeled products (5 μL each) were pooled with 2 μL basic fuchsin stop/loading buffer (LI-COR). Microsatellites could be amplified in both single and multiplex reactions. For multiplex PCR, a primer pair labeled with IRDye700 and a primer pair with IRDye800, both with the same annealing temperature, were combined in the same 10-μL reaction. Loading buffer (2 μL) was added at the completion of the reaction.

For both methods, gel electrophoresis and visualization of the microsatellite alleles were performed using the IR² Gene Read IR 4200 DNA Analyzer, which detects fluororescently labeled PCR products in real time. Gels were 25 cm in length and 0.25 cm in thickness and contained 7% Long Ranger™ gel solution (BMA, Rockland, ME, USA) and 7 M urea in 1×TBE buffer (89 mM Tris, pH 8.3, 89 mM boric acid, and 2 mM EDTA). A 64-well, square-tooth comb was used. One microliter of the reaction mixture was loaded into the wells with an 8-channel sample-loading syringe (Hamilton, Reno, NV, USA), and 0.5 μL each IRDye700 and 800-labeled sizing standard ranging from 50 to 350 bp were loaded twice per gel. The samples and sizing standard were heated for 2 min at 95°C and kept on ice before gel loading. The gels were run in 1×TBE buffer with the electrophoresis conditions defaulted as voltage at 1500 V, current at 40 mA, power at 40 W, signal channel at 3, scan speed at 3, and 16-bit pixel depth for the collection of TIFF image files. The plate temperature was maintained at 48°C. The gels were run for 1.5–2.5 h, depending on the expected allele size. The infrared fluorescently labeled PCR product was detected with the automated DNA analyzer and stored as the autoradiogram-like TIFF images (Figure 1). Each gel was used twice. Assessment was based on 9920 datapoints for each method.

The mating type configuration and expected PCR products were consistent between the standard PAGE gels and the direct fluorescent primer method. The latter method did not show addi-

Benchmarks

tional or missing alleles. Missing PCR products were 5%–10%, a rate similar to PAGE gels. The re-optimization of annealing temperatures was important for adapting this method to *P. taeda* microsatellites. For example, the marker PtTX3106 did not amplify initially for the fluorescently labeled forward primer PCR at the annealing temperature of 65°C. A temperature of 55°C produced a PCR profile optimized for the DNA analyzer (Figure 1).

The M13-tailed method showed inconsistent performance for *P. taeda* microsatellites. Of the 80 M13-tailed primer sets that showed amplification, 34 primer sets (42.5%) showed the addition or loss of at least one allele in comparison to the standard PAGE gel. This is a conservative estimate of failure because only those microsatellites that amplified with the M13-tailed primer present could be compared. Adding false alleles was more common than losing actual alleles; false alleles were added in 20 of the 34 cases. The mating-type configuration switched from BC to IC in 14 out of 20 of the cases where actual alleles were added. Fourteen of the 34 cases lost one or more alleles. Other banding artifacts were common, resulting in

blurred gel images, and lanes missing PCR products altogether were common (Figure 1). Our 1:2 molar ratio of labeled M13 primer to the M13-tailed primer is common for large plant genomes using infrared detection. The molar ratios of 1:15 and 1:18 used for mammalian genomes (1,12) have not been tested with conifer genomes.

We hypothesize that the inconsistency of the M13-tailed primer method is due to two attributes of the *P. taeda* genome, its large genome size and high degree of duplication. The forward primer with its 19-bp M13 conjugate caused a high degree of nonspecific priming. Nonspecific priming included self-dimerization, primer pair dimers, and primer hairpins. The direct fluorescent primer method was highly consistent and, thus, more cost effective for *P. taeda* microsatellites. Direct fluorescent primer labeling did require the optimization of re-annealing temperatures because the IRDye is a large molecule (12) and may cause steric hindrance during direct incorporation, thus interfering with the binding of labeled primers to the template during PCR. Automated genotyping results for large plant genomes should be tested against

PAGE gels to ensure high-quality marker data collection.

In summary, a three-primer or M13-tailed primer protocol was compared with direct fluorescent labeling for high-throughput data collection of *P. taeda* microsatellites using the IR² 4200 DNA Analyzer. Standard PAGE gels were used as a control. Fluorescently labeled primer methods gave results that were consistent with PAGE gels for *P. taeda* nuclear microsatellites. The resulting PCR products had the expected product size and inheritance patterns. By contrast, the M13-tailed primer method was highly inconsistent. Of the 80 M13-tailed primer sets that showed PCR amplification, 34 primer sets (42.5%) showed the addition or loss of one or more alleles, in comparison to the standard PAGE gel and gels based on PCR products with direct fluorescent labeling. For the M13-tailed primer method, adding false alleles was more common than losing actual alleles. We hypothesize that the M13-tailed primer method performs poorly with large, highly repetitive plant genomes. Large and small plant genomes have not yet been compared for the M13-tailed versus fluorescently labeled primers.

Figure 1. A comparison of PtTX3106 for the M13-tailed primer method (A) and the direct fluorescent primer method (B). Lanes 1–11 are progeny segregating for four alleles. For the M13-tailed primer method, alleles are missing, and there is an absence of PCR amplification for lanes 1, 2, 5, and 6.

REFERENCES

1. Boutin-Ganache, L., M. Raposo, M. Raymond, and C.F. Deschepper. 2001. M13-tailed primers improve the readability and usability of microsatellite analyses performed with two different allele-sizing methods. *BioTechniques* 31:24–28.
2. Devey, M.E., T. Fiddler, B. Liu, S. Knapp, and D.B. Neale. 1994. An RFLP map for loblolly pine based on a three-generation outbred pedigree. *Theor. Appl. Genet.* 88:273–278.
3. Doyle, J.J. and J.L. Doyle. 1987. A rapid DNA isolation procedure for small quantities of fresh leaf tissue. *Phytochem. Bull.* 19:11–15.
4. Echt, C.S., P. May-Marquardt, M. Hsieh, and R. Zahorchak. 1996. Characterization of microsatellite markers in eastern white pine. *Genome* 39:1102–1108.
5. Elsik, C.G., V.T. Minihan, S.E. Hall, A.M. Scarpa, and C.G. Williams. 2000. Low-copy microsatellite markers for *Pinus taeda* L. *Genome* 43:550–555.
6. Elsik, C.G. and C.G. Williams. 2000. Retroelements contribute to the excess low-copy DNA in pine. *Mol. Gen. Genet.* 264:47–55.
7. Elsik, C.G. and C.G. Williams. 2001. Families of clustered microsatellites in a conifer genome. *Mol. Gen. Genet.* 265:535–542.

Benchmarks

8. Joyner, K.L., X.-R. Wang, J.S. Johnston, H.J. Price, and C.G. Williams. 2001. DNA content for Asian pines parallels New World relatives. *Can. J. Bot.* 79:179-191.
9. Kinlaw, C.S. and D.B. Neale. 1997. Complex gene families in the pine genome. *Trends Plant Sci.* 2:356-359.
10. Kriebel, H.B. 1985. DNA sequence components of the *Pinus strobus* nuclear genome. *Can. J. For. Res.* 15:1-4.
11. Oetting, W.S., H.K. Lee, D.J. Flanders, G.L. Wiesner, T.A. Sellers, and R.A. King. 1995. Linkage analysis with multiplexed short tandem repeat polymorphisms using infrared fluorescence and M13 tailed primers. *Genomics* 30:450-458.
12. Roy, R., D.L. Steffens, B. Gartside, G.Y. Jang, and J.A. Brumbaugh. 1996. Producing STR locus patterns from bloodstains and other forensic samples using an infrared fluorescent automated DNA sequencer. *J. Forensic Sci.* 41:418-424.
13. Schwengel, D.A., A.E. Jedlicka, E.L. Nanthakumar, J.L. Weber, and R.C. Levitt. 1994. Comparison of fluorescence-based semi-automated genotyping of multiple microsatellite loci with autoradiographic techniques. *Genomics* 22:46-54.
14. Steffens, D.L., S.L. Sutter, and S.C. Roemer. 1993. An alternate universal forward primer for improved automated DNA sequencing of M13. *BioTechniques* 15:580-581.
15. Williams, C.G. 1998. QTL mapping in outbred pedigree, p. 81-94. In A.H. Paterson (Ed.), *Molecular Dissection of Complex Traits*. CRC Press, Boca Raton, FL.
16. Zhou, Y., T. Bui, L.D. Auckland, and C.G. Williams. Undermethylated DNA as a source of microsatellites from a conifer genome. *Genome* (In Press.)

This work was supported by Department of Energy grant no. DE-FC07-00ID13877, National Science Foundation Bioinstrumentation grant no. NSF DBI-9871390, and Texas Higher Education Coordinating Board grant no. 999902-090. The authors thank Dr. Patricia Klein, Jacques Obert, and Natalie Unruh for technical assistance with the LI-COR automated DNA analyzer at the Borlaug Center for Southern Crop Improvement. Address correspondence to Dr. Claire G. Williams, Faculty of Genetics and Department of Forest Science, Texas A&M University, College Station, TX 77843-2135, USA. e-mail: claire-williams@tamu.edu

Received 1 August 2001; accepted 18 September 2001.

Yi Zhou, Thomas Bui, Lisa D. Auckland, and Claire G. Williams
*Texas A&M University
College Station, TX, USA*

Kit-Based, Low-Toxicity Method for Extracting and Purifying Fungal DNA from Ectomycorrhizal Roots

BioTechniques 32:52-56 (January 2002)

Ectomycorrhizal fungi are ecologically and economically important because when they form associations with roots of species such as pines, oaks, and eucalypts, they may increase nutrient and water uptake (15). Functional and taxonomic diversity among the ectomycorrhizal fungi is great and thus, in many studies, it is important to identify individual fungal species. During the past few years, researchers have perfected DNA-based methods for the identification of ectomycorrhizal fungi. In most cases, this involves the extraction and purification of DNA from colonized roots, amplification of fungal genomic DNA, and either analysis of restriction fragments or sequencing (1, 3-5,7,9,10).

The extraction of DNA from ectomycorrhizal roots, as commonly published, involves some combination of manual grinding, freezing and thawing, and incubation at elevated temperatures, while purification may involve chloroform or phenol/chloroform extraction of impurities and precipitation of DNA by alcohol (3,5,7,10). The process may take hours to complete, and it requires the use of a fume hood to vent chloroform and phenol vapors. We desired a simple, inexpensive, rapid, and less hazardous method for extracting and purifying readily amplifiable fungal DNA from ectomycorrhizal roots, and a kit-based method seemed ideal. However, as far as we are aware, there are no kits specifically designed to extract and purify fungal DNA from ectomycorrhizal roots. Here, we report on the successful use of a commercially available kit to do this without manual grinding, specialized shaking apparatus, or phenol/chloroform.

The three extraction/purification methods reported here use the UltraClean™ Microbial DNA Kit (Mo Bio Laboratories, Solana Beach, CA, USA).

The kit was originally designed to extract DNA from solution-cultured microbial cell pellets. It consists of 1.9-mL microcentrifuge extraction tubes containing garnet beads, MicroBead buffer solution, a detergent-based extraction solution (M1), an acetate-based solution to precipitate impurities (M2), a salt solution to cause the DNA to bind to the supplied silica column (M3), an ethanol washing solution (M4), and 10 mM Tris (M5) for DNA elution.

For Method 1, clusters of ectomycorrhizal roots comprising 2-4 root tips were sampled by hand from the fermentation layer of the forest floor from an approximately 65-year-old red pine (*Pinus resinosa*) plantation located in State College, Centre County, PA, USA. The ectomycorrhizal root clusters were viewed using a dissecting microscope (10-40×) with reflected light to separate them according to morphology. The clusters of a single morphotype, morphotype G (unpublished data), were selected. All root clusters were frozen at -20°C before use. Two days later, DNA from two replicate ectomycorrhizal root clusters was extracted and purified. Each root cluster was placed in its own extraction tube. Then, 300 µL MicroBead solution were added, followed by 50 µL M1 and 50 µL inhibitor removal solution (IRS) (Mo Bio Laboratories). IRS is included in the UltraClean Soil DNA Kit to inactivate the phenolic components of soil organic matter that can inhibit DNA amplification (16). Phenolic compounds are also present in the roots of many ectomycorrhizal trees (8,13,14, 17), so it was used here. Each extraction tube was heated to 65°C for 15 min in an oven to aid in the extraction. The tubes were then clipped to the vortex adapter (Mo Bio Laboratories) attached to a Vortex Genie 2 Mixer® (Fisher Scientific, Pittsburgh, PA, USA) and agitated at the highest speed for 15 min. The adapter consists of a 14.5-cm diameter plastic plate with 12 plastic clips around the perimeter to hold the tubes. We attached the adapter to the mixer with a strong adhesive, not with the loop and pile tape supplied by the manufacturer. This transferred more of the energy from the mixer to the tubes. A newer version of the adapter that does not require this adhesive step is