of potential inhibitors such as hemoglobin. This buffy-coat direct PCR can be performed on either citrated or EDTA-anticoagulated blood. Citrated blood was of particular interest since the same blood sample could be used for both plasma phenotypic and genotypic analyses of APC-resistance.

This rapid buffy-coat direct PCR allowed for amplification using a simpler protocol than that of Bertina et al. (3). It is a very simple procedure because it requires neither DNA extraction nor pre-isolation of white blood cells, and it is well-suited for routine diagnostic work.

REFERENCES

1. Ausubel, F.M., R. Brent, R.E. Kingston, D.D. Moore, J.G. Seidman, J.A. Smith and K. Struhl. 1992. Preparation and analysis of data, p.2.2.1-2.2.2. Current Protocols in Molecular Biology. John Wiley & Sons, New York.
2. Balnaves, M.E., S. Nasioulas, H.H.M. Dahl and S. Forrest. 1991. Direct PCR from CVS and blood lysates for detection of cystic fibrosis and Duchenne muscular dystrophy deletions. *Nucleic Acids Res.* 19:1155.
3. Bertina, R.M., B.P.C. Koeleman, T. Koster, F.R. Rosendaal, R.J. Dirven, H. de Ronde, P.A. van der Veiden and P.H. Reitsma. 1994. Mutation in blood coagulation factor V associated with resistance to activated protein C. *Nature* 369:64-67.
4. Cobb, B.D. and J.M. Clarkson. 1994. A simple procedure for optimising the polymerase chain reaction (PCR) using modified Taguchi methods. *Nucleic Acids Res.* 22:3801-3805.
5. Dahlbäck, B. 1995. Resistance to activated protein C, the Arg⁵⁰⁶ to Gln mutation in the factor V gene, and venous thrombosis. *Thromb. Haemost.* 73:739-742.
6. Dahlbäck, B., M. Carlsson and P.J. Svensson. 1993. Familial thrombophilia due to a previously unrecognized mechanism characterized by poor anticoagulant response to activated protein C: prediction of a cofactor to activated protein C. *Proc. Natl. Acad. Sci. USA* 90:1004-1008.
7. Greengard, J.S., X. Sun, X. Xu, J.A. Fernandez, J.H. Griffin and B. Evatt. 1994. Activated protein C resistance caused by Arg506Gln mutation in factor Va. *Lancet* 343:1362-1363.
8. Koeleman B.P.C., P.H. Reitsma, C.F. Allaart and R.M. Bertina. 1994. Activated protein C resistance as an additional risk factor for thrombosis in protein C-deficient families. *Blood* 84:1031-1035.
9. Lewin, H.A. and J.A. Stewart-Haynes. 1992. A simple method for DNA extraction from leukocytes for use in PCR. *BioTechniques* 13:522-524.
10. McHale, R.H., P.M. Stapleton and P.L. Bergquist. 1991. Rapid preparation of blood and tissue samples for polymerase chain reaction. *BioTechniques* 10:20-23.
11. Mercier, B., C. Gaucher, O. Feugas and C. Mazurier. 1990. Direct PCR from whole blood, without DNA extraction. *Nucleic Acids Res.* 18:5908.
12. Nordvåg, B.-Y., G. Husby and M. Raafat El-Gewely. 1992. Direct PCR of washed blood cells. *BioTechniques* 12:490-493.
13. Svensson, P.J. and B. Dahlbäck. 1994. Resistance to activated protein C as a basis for venous thrombosis. *N. Engl. J. Med.* 330:517-522.
14. Zöller, B. and B. Dahlbäck. 1994. Linkage between inherited resistance to activated protein C and factor V gene mutation in venous thrombosis. *Lancet* 343:1536-1538.
15. Zöller, B., P.J. Svensson, X. He and B. Dahlbäck. 1994. Identification of the same factor V gene mutation in 47 out of 50 thrombosis-prone families with inherited resistance to activated protein C. *J. Clin. Invest.* 94:2521-2524.

We thank A. Besson and C. Coudoux for technical assistance. Address correspondence to Gilles Pernod, Laboratoire d'Hématologie, Hopital A Michalon, CHU de Grenoble, BP 217, 38043 Grenoble Cédex 9, France.

Received 14 February 1996; accepted 11 November 1996.

Gilles Pernod, Pascal Mossuz and Benoit Polack
CHU de Grenoble
Grenoble, France

Spectrophotometric Determination of Oxidative Metabolism

BioTechniques 22:841-844 (May 1997)

Oxidative metabolism is a good indicator of the activation state of several cell types including macrophage(s) (Mφ), eosinophils and B cells (1,2,5,6). This indicator has been successfully used to monitor the activation state of Mφ and remains an important technique for experimental assessment of this complex process (2,7). Traditionally, oxidative metabolism has been assayed by the conversion of nitroblue tetrazolium (NBT) to blue/black formazan deposits in the target cells and the subsequent microscopic examina-

tion and enumeration of cells containing these deposits (2,3,4). This method is very labor-intensive and potentially subject to investigator bias. We have developed a method for examining this conversion spectrophotometrically, thereby eliminating observer subjectivity and allowing a more accurate determination in a fraction of the time.

Initial experiments were conducted to compare the spectrophotometric assay with the traditional microscopic method. Mφ are plated at three different concentrations in a 96-well plate and examined both microscopically and spectrophotometrically. The procedure is carried out under sterile conditions by plating cells at $1-5 \times 10^4$ cells per well in 160 μL and adding 40 μL of NBT to each well. NBT is purchased from Sigma Chemical (St. Louis, MO, USA) as a lyophilized product in the presence of phosphate buffer and NaCl. It is resuspended in sterile H₂O at 2.0 mg/mL. The unconverted NBT has an absorbance spectrum that peaks at 385 nm and falls to zero after 490 nm (Figure 1A). Once converted to formazan by the addition of NaOH, the absorbance spectrum varies only marginally, exhibiting a peak absorbance at 655 nm, well beyond the peak for the unconverted form (Figure 1B). Thus, no assay interference is demonstrated at wavelengths 490-700 nm. Additionally, since we artificially converted the NBT to formazan through the addition of NaOH to determine its peak absorbance, we examined the effect of altering pH on the absorbance at 655 nm. The changes in optical density (OD)₆₅₅ did not vary as a function of pH changes within the range of pH 2.97-11.66 (data not shown). Likewise, the contribution of the cells themselves to the absorbance at 655 nm at each of the cell numbers examined was negligible (data not shown). Following the NBT addition, the plate is shaken gently for 1 min and incubated for 24 h in a humidified chamber at 37°C and 7% CO₂. After 24 h, microscopic examination is performed using a Diaphot-TMD Phase Contrast Microscope (Nikon, Tokyo, Japan) at 100× final magnification, and the number of positive cells per field is determined. The plates are then shaken vigorously for 5 s, read on a SPECTRAMax™ 250 Plate Reader

Benchmarks

(Molecular Devices, Menlo Park, CA, USA) at 655 nm and analyzed using SOFTmax® PRO 1.1 for Macintosh® (Molecular Devices).

Figure 2 shows a representative experiment in which RAW 264.7 cells, a murine Mφ cell line, were plated at 1×10^4 , 2.5×10^4 and 5×10^4 cells per well in triplicate in RPMI-1640 containing 10% fetal bovine serum (FBS) (Life Technologies, Gaithersburg, MD, USA). NBT was added, and cultures were incubated for 24 h at 37°C in 7% CO₂. Medium alone with NBT added was included as a blank. At the end of the incubation period, the plates were examined microscopically at 100× magnification to determine the number of positive cells per microscopic field, and counts from three fields in the center of the wells were averaged for each well (Figure 2, ■). The absorbance of the wells at 655 nm was then determined using the SPECTRAMax 250 Plate Reader (Figure 2, ●). As shown, the spectrophotometric and microscopic examinations both yielded a similar linear relationship between the quantitated level of formazan deposition and cell number.

An additional observation was made regarding optimal cell density. In analyzing the visual count at various cell concentrations, we noted that at lower concentrations (below 1×10^4 cells per well), the number of positive cells per microscopic field did not exhibit a linear correlation with the number of cells per well. In fact at 1×10^4 cells per well, a slightly decreased correlation can be noted in Figure 2 (■). This suggests that at extremely low cell concentrations, the conversion of NBT to formazan may not accurately yield a quantitative assessment of oxidative metabolism. It is therefore advisable to optimize the cellular concentration for each cell type to be assayed. However, cell numbers of 2.5 – 5×10^4 cells per well represent a widely applicable range that is linear for all cell lines tested in these studies.

Although this protocol gave reproducible results, we were concerned that uneven deposition of formazan within the individual cells might affect the spectrophotometric reading even though the plates were shaken vigorously by the SPECTRAMax 250 imme-

diately before absorbance determinations. Therefore, several agents that might allow for cellular disruption and formazan deposit distribution were examined. Treatment of the wells with a solution of 0.5% sodium dodecyl sulfate (SDS) in 0.35% (vol/vol) glacial acetic acid reproducibly effected the desired cellular disruption without interfering with the spectrophotometric

assay (Figure 2, △). This solution did not cause NBT reduction in the absence of cells (data not shown). Other agents were examined with varying degrees of success. Bubbles were often observed in the wells when using higher concentrations of SDS with or without acetic acid, thus causing errant OD₆₅₅ values. Chloroform by itself converted NBT, and phosphoric acid consistently low-

Figure 1. Determination of the absorbance spectrum of NBT and formazan deposits. Lyophilized, unconverted NBT was reconstituted with sterile water to a concentration of 2 mg/mL and examined for absorbance at wavelengths 380–700 nm (A). The NBT was then converted to formazan deposits by the addition of NaOH and examined for absorbance over the same range of wavelengths (B).

Figure 2. Comparison of levels of NBT activity assessed by visual count vs. OD₆₅₅ determination in RAW 264.7 cells. RAW 264.7 cells were plated in triplicate in sterile 96-well plates at 1×10^4 , 2.5×10^4 and 5×10^4 cells per well. NBT was added to a final concentration of 400 mg/mL, and the cells were cultured for 24 h at 37°C. The wells were then examined on the Phase Contrast Microscope, and the number of cells containing blue/black formazan deposits was enumerated as described. The number of positive cells per 100× microscope field are shown $\pm$ standard error of the mean (SEM) (■-). The same wells were then read on a SPECTRAMax 250 96-well Plate Spectrophotometer at 655 nm, and the absorbance $\pm$ SEM was determined (●-). Finally, SDS-acetic acid solution was added to disrupt and resuspend the samples, and the OD₆₅₅ $\pm$ SEM is shown (△-).

Benchmarks

ered the OD₆₅₅, perhaps as a result of formazan dissolution (data not shown).

In addition to the experiment outlined above, we have examined interferon- γ (IFN- γ) and phorbol ester activation of RAW 264.7, WEHI-3, WEHI-3 D⁻ and WEHI-3 D⁺ murine monocytic cell lines. In all cases, cellular activation was accompanied by an increased conversion of NBT as determined by an increased OD₆₅₅. While not all cell lines were examined both microscopically and spectrophotometrically under all combinations of culture treatments, all untreated lines were examined by both methods, and at least two cell lines (the RAW 264.7 and WEHI-3 cell lines) were examined by both methods for each treatment group (data not shown). The data collected in these experiments indicate that the spectrophotometric method of determining NBT conversion is applicable to numerous cell lines.

REFERENCES

1. Kobayashi, S., S. Imajoh-Ohmi, F. Kuribayashi, H. Nunoi, M. Nakamura and S. Kanegasaki. 1995. Characterization of the superoxide-generating system in human peripheral lymphocytes and cell lines. *J. Biochem.* 117:758-765.
2. Li, J., I. King and A.C. Sartorelli. 1994. Differentiation of WEHI-3B D⁺ myelomonocytic leukemia cells induced by ectopic expression of the protooncogene *c-jun*. *Cell Growth Differ.* 5:743-751.
3. Park, B.H., W.D. Bigger, P. L'Esperance and R.A. Good. 1972. N.B.T. test on monocytes of neutropenic patients. *Lancet* 1:1064.
4. Park, B.H., S.M. Fikrig and E.M. Smithwick. 1968. Infection and nitroblue-tetrazolium reduction by neutrophils: a diagnostic aid. *Lancet* 2:532-534.
5. Porter, C.D., M.H. Parker, M.K. Collins, R.J. Levinsky and C. Kinnon. 1992. Superoxide production by normal and chronic granulomatous disease (CGD) patient-derived EBV-transformed B cell lines measured by chemiluminescence-based assays. *J. Immunol. Methods* 155:151-157.
6. Thomas, E.L., P.M. Bozeman, M.M. Jefferson and C.C. King. 1995. Oxidation of bromide by human leukocyte enzymes myeloperoxidase and eosinophil peroxidase. Formation of bromamines. *J. Biol. Chem.* 270:2906-2913.
7. Yoshida, M., T. Eguchi, N. Ikekawa and N. Saijo. 1995. Inhibition of vitamin D₃-induced cell differentiation by interferon- γ in HL-60 cells determined by a nitroblue tetrazolium reduction test. *J. Interferon Cyto. Res.* 15:965-971.

Address correspondence to Donna M. Paulnock, Department of Medical Microbiology and Immunology, University of Wisconsin Medical School, 1300 University Avenue, Madison, WI 53706, USA. Internet: paulnock@facstaff.wisc.edu

Received 6 May 1996; accepted 6 November 1996.

Mary A. Lokuta, Glen H. Mehring and Donna M. Paulnock
*University of Wisconsin Medical School
Madison, WI, USA*