Use of the Precellys® Evolution homogenizer for unbiased DNA extraction from the ZymoBIOMICS™ microbial community standard

Zymo Research Corp (Irvine, CA, USA) and Bertin Instruments (Rockville, MD, USA)

Microbiome profiling via Next Generation Sequencing (NGS) techniques is rapidly changing the landscape of microbiology. However, to ensure accurate microbiome measurements, it is essential to ensure the microbiome workflow is unbiased and robust [1]. For instance, it has been reported that bias in microbiomics analyses can be introduced at every step [2-4] including the DNA extraction step if microbial cells are not uniformly and completely lysed. Mechanical cell lysis using beads of various composition (bead beating) is a generally accepted method used to lyse various microbes [5] and, although a variety of mechanical lysis devices and instruments are available, their performance greatly varies [6-8]. Until recently, there has not been a convenient way to assess the efficiency of lysis afforded by any particular instrument, however, with the release of the ZymoBIOMICS™ Microbial Community Standard (Zymo Research Corp, Irvine, CA, USA) this is now possible. The ZymoBIOMICS Microbial Community is a mock microbial community standard comprised of ten well-characterized organisms (i.e., 5 gram-positive, 3 gram-negative, and 2 yeast) that are mixed at defined proportions. The standard contains microbes of different cell wall recalcitrance and of different cell size making it ideal for evaluating cell lysis efficiency. In this study, the efficiency of microbial lysis with the ZymoBIOMICS Microbial Community Standard was assessed using the Precellys® Evolution tissue homogenizer combined with Cryolys® Evolution cooling unit (Bertin Instruments, Rockville, MD, USA).

Materials

- Instrument: Precellys Evolution and Cryolys Evolution (Bertin Instruments)
- Sample: DNA of the twelve aliquots (75 µl each) of the ZymoBIOMICS Microbial Community Standard extracted using the ZymoBIOMICS DNA Miniprep Kit (Zymo Research Corp.) following the manufacturer's protocol.
- Buffer: 725 µl of Lysis Solution per sample included in the ZymoBIOMICS DNA Miniprep Kit (Zymo Research Corp.)

Protocols

- Cell lysis was performed using a Precellys Evolution set at maximum speed (10,000 rpm) fitted with a Cryolys Evolution cooling unit set at 4°C.
- Lysis of the standard was assessed in triplicate and several run where tested in order to validate the optimal run:
 - 1*60 sec (total time 1 min)
 - 3*60 sec, 60 sec break after each minute (total time 5 min)
 - 5*60 sec, 60 sec break after each minute (total time 9 min)
- Extracted DNA was quantified using a Qubit™ dsDNA BR Assay Kit (Thermo Fisher Scientific) and the microbial composition of the extracted DNA determined via 16S rRNA gene sequencing using an Illumina® MiSeq™ (2x300bp).
- Sequencing and data analysis were performed by the ZymoBIOMICS Sequencing Service (Zymo Research).
- The control cell lysis was performed as described in the cell lysis section using a Supplier X tissue homogenizer set at maximum speed (6.5m/s).
- Lysis of the control was optimized by using the following run: 5*60 sec, 5 minutes break after each minute (total time 25 min)
- DNA extraction and sequencing were performed as explained in the Cell lysis protocol section.

Results

DNA quantification indicated the yields of extracted DNA reached a maximum (compared to control) when lysed for 1 minute using the Precellys Evolution tissue homogenizer (Figure 3). This indicates that efficient cell lysis can be achieved by homogenizing cells for 1 minute instead of 5 minutes which was previously reported. These data were supported by 16S sequencing data that showed the microbial composition reflected the theoretical closely indicating that complete lysis was achieved in as little as 1 minute of bead beating (Figure 4). In addition, standard composition is consistent and stays stable with the theoretical value even after 5 minutes of homogenization using Precellys Evolution tissue homogenizer.

Figure 1. Precellys® Evolution homogenizer & Cryolys® Evolution cooling system.

Figure 2. Precellys® protocol on the Human Machine Interface.

Figure 3. DNA concentrations achieved after homogenizing cells for 1, 3 and 5 min using the Precellys® Evolution tissue homogenizer representing a total processing time of 1, 5 and 9 min respectively. High yield DNA was obtained in as little as 1-minute cell lysis. The 'Control' sample is the DNA yield obtained using an alternative homogenizer (Supplier X) and was achieved in approximately 5 min' bead beating time and 25 min processing time in total.

Figure 4. Microbial composition analysis (16S) closely matches the ZymoBIOMICS™ Microbial Community Standard's theoretical composition in as little as 1 minute using the Precellys® Evolution tissue homogenizer. DNA was extracted from ZymoBIOMICS Microbial Community Standard using the ZymoBIOMICS DNA Miniprep Kit and Precellys Evolution tissue homogenizer and yields determined at the processing times indicated.

Conclusions

The Precellys Evolution tissue homogenizer is capable of achieving complete, unbiased cell lysis in only 1 minute as assessed using the ZymoBIOMICS Microbial Community Standard in conjunction with the ZymoBIOMICS DNA Miniprep Kit. The optimized microbial lysis workflow reported here, serves as a validated procedure for accurate downstream microbiome measurements including amplicon sequencing, shotgun sequencing, RNA-based sequencing procedures. Not only does the workflow provide for efficient uniform microbial lysis and higher yields of inhibitor-free DNA, but it also facilitates reduced processing times with the entire extraction procedure requiring less than 20 minutes.

About the Precellys Evolution

Based on a decade of sample preparation expertise, Bertin designed the Precellys Evolution especially for L3 laboratories. With its optimal robustness, safety, efficiency and flexibility, the instrument is the universal tissue homogenizer, perfect for either routine or R&D projects.

For sensitive sample preparation, the tissue homogenizer can be combined with the Cryolys Evolution. This cooling unit can ensure the integrity of specific molecules, such as proteins or RNA, guarantying a better efficiency and sensitivity for downstream analysis. The Cryolys Evolution is designed to fit on the top of the Precellys Evolution homogenizer.

References

1. Sinha, Rashmi *et al.* The microbiome quality control project: baseline study design and future directions. *Genome Biol.* 16.1 (2015): 276.
2. Song, Se Jin *et al.* Preservation methods differ in fecal microbiome stability, affecting suitability for field studies. *mSystems* 1.3 (2016): e00021-16.
3. Gerasimidis, Konstantinos *et al.* The effect of DNA extraction methodology on gut microbiota research applications. *BMC research notes* 9.1 (2016): 365.
4. Jones, Marcus B *et al.* Library preparation methodology can influence genomic and functional predictions in human microbiome research. *Proc. Natl Acad. Sci. USA* 112.45 (2015): 14024-14029.
5. Vishnivetskaya, Tatiana A *et al.* Commercial DNA extraction kits impact observed microbial community composition in permafrost samples. *FEMS Microbiology Ecology* 87.1 (2014): 217-230.
6. Wesolowska-Andersen, Agata *et al.* Choice of bacterial DNA extraction method from fecal material influences community structure as evaluated by metagenomic analysis. *Microbiome* 2.1 (2014): 19.
7. Hsieh, Yu-Hsin *et al.* Impact of different fecal processing methods on assessments of bacterial diversity in the human intestine. *Frontiers in Microbiology* 7 (2016).
8. Kennedy, Nicholas A *et al.* The impact of different DNA extraction kits and laboratories upon the assessment of human gut microbiota composition by 16S rRNA gene sequencing. *PloS ONE* 9.2 (2014): e88982.